


ISM02.4

LB 87694
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44057



A= 0.0000e+000

B= 2.2683e-004

C= -1.0392e-002

Rho= 0.9999893

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.002	0.002	53	0.000	53					-
0.2	0.200	0.181	-0.019	844	0.0 %	844					-10
2.5	2.500	2.522	0.022	11166	0.0 %	11166					1
5.0	5.000	4.989	-0.011	22042	0.0 %	22042					0
7.5	7.500	7.522	0.022	33207	0.0 %	33207					0
10.0	10.000	9.983	-0.017	44057	0.0 %	44057					0

NB
5/25/17

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	53	-		0 ISM02.4	PPB	5/25/2017 11:20	Std
	0.2 S0.2	844	-		0.2 ISM02.4	PPB	5/25/2017 11:22	Std
	2.5 S2.5	11166	-		2.5 ISM02.4	PPB	5/25/2017 11:24	Std
	5 S5	22042	-		5 ISM02.4	PPB	5/25/2017 11:26	Std
	7.5 S7.5	33207	-		7.5 ISM02.4	PPB	5/25/2017 11:29	Std
	10 S10	44057	-		10 ISM02.4	PPB	5/25/2017 11:31	Std
ICV	ICV	16942	3.8326 -		ISM02.4	PPB	5/25/2017 11:34	SMPL
ICB	ICB	-25	-0.0161 -		ISM02.4	PPB	5/25/2017 11:36	SMPL
CCV016	CCV016	21745	4.9221 -		ISM02.4	PPB	5/25/2017 11:38	SMPL
CCB016	CCB016	-21	-0.0152 -		ISM02.4	PPB	5/25/2017 11:41	SMPL
PB99269BL	PBS004	-9	-0.0124 -		ISM02.4	PPB	5/25/2017 11:43	SMPL
I3213-19	MX2999	32779	7.425 -		ISM02.4	PPB	5/25/2017 11:45	SMPL
I3213-20	MX2999D	31951	7.2372 -		ISM02.4	PPB	5/25/2017 11:47	SMPL
I3213-21	MX2999S	42472	9.6237 -		ISM02.4	PPB	5/25/2017 11:52	SMPL
CCV017	CCV017	21317	4.825 -		ISM02.4	PPB	5/25/2017 11:55	SMPL
CCB017	CCB017	-26	-0.0163 -		ISM02.4	PPB	5/25/2017 11:57	SMPL