


7471A

LB87695
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42636



A= 0.0000e+000

B= 2.3476e-004

C= 9.4016e-003

Rho= 0.9999921

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	20	0.000	20				
0.2	0.200	0.198	-0.002	804	0.0 %	804				
2.5	2.500	2.489	-0.011	10561	0.0 %	10561				
5.0	5.000	5.005	0.005	21279	0.0 %	21279				
7.5	7.500	7.476	-0.024	31804	0.0 %	31804				
10.0	10.000	10.019	0.019	42636	0.0 %	42636				

LB87695 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 25 May 2017 12:46:46

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	20	-	0.00007471A	PPB	25 May 2017 12:49:29	S	Std
0.2 - 1	50.2	804	-	0.20007471A	PPB	25 May 2017 12:51:47	S	Std
2.5 - 1	52.5	10561	-	2.50007471A	PPB	25 May 2017 12:54:04	S	Std
5.0 - 1	55	21279	-	5.00007471A	PPB	25 May 2017 12:56:22	S	Std
7.5 - 1	57.5	31804	-	7.50007471A	PPB	25 May 2017 12:58:39	S	Std
10.0 - 1	510	42636	-	10.00007471A	PPB	25 May 2017 13:00:56	S	Std
ICV74 - 1	ICV74	16382	3.8552	-7471A	PPB	25 May 2017 13:03:34	U	SMPL
ICB74 - 1	ICB74	25	0.0153	-7471A	PPB	25 May 2017 13:05:49	U	SMPL
CCV18 - 1	CCV18	21113	4.9659	-7471A	PPB	25 May 2017 13:08:05	U	SMPL
CCB18 - 1	CCB18	12	0.0122	-7471A	PPB	25 May 2017 13:10:21	U	SMPL
CRA - 1	CRA	842	0.2071	-7471A	PPB	25 May 2017 13:12:36	U	SMPL
HighStd - 1	HighStd	42276	9.9341	-7471A	PPB	25 May 2017 13:14:52	U	SMPL
ChkStd - 1	ChkStd	28985	6.8139	-7471A	PPB	25 May 2017 13:17:08	U	SMPL
PB99270BL - 1	PBS	-53	-0.0030	-7471A	PPB	25 May 2017 13:19:25	U	SMPL
I3329-01 - 1	B-15(3-9)	13286	3.1284	-7471A	PPB	25 May 2017 13:23:59	U	SMPL
I3329-02 - 1	B-14(6-9)	19943	4.6912	-7471A	PPB	25 May 2017 13:26:16	U	SMPL
I3329-03 - 1	B-13(5-8)	141060	33.1245	-7471A	PPB	25 May 2017 13:28:34	U	SMPL
I3329-04 - 1	B-3(0-3)	5142	1.2165	-7471A	PPB	25 May 2017 13:30:51	U	SMPL
I3329-05 - 1	B-11(0-3)	18745	4.4100	-7471A	PPB	25 May 2017 13:33:09	U	SMPL
CCV19 - 1	CCV19	21287	5.0067	-7471A	PPB	25 May 2017 13:35:25	U	SMPL
CCB19 - 1	CCB19	-33	0.0017	-7471A	PPB	25 May 2017 13:37:41	U	SMPL
I3329-06 - 1	B-7(3-7)	1378	0.3329	-7471A	PPB	25 May 2017 13:39:57	U	SMPL
I3329-07 - 1	B-8(0-3)	36452	8.5668	-7471A	PPB	25 May 2017 13:42:12	U	SMPL
I3329-08 - 1	B-18(0-5)	282969	66.4388	-7471A	PPB	25 May 2017 13:44:28	U	SMPL
I3329-09 - 1	B-17(0-5)	28460	6.6906	-7471A	PPB	25 May 2017 13:46:45	U	SMPL
I3329-10 - 1	B-9(2-6)	16391	3.8573	-7471A	PPB	25 May 2017 13:49:04	U	SMPL
I3331-02 - 1	SB-22COMP	1353	0.3270	-7471A	PPB	25 May 2017 13:51:21	U	SMPL
I3331-04 - 1	SB-23COMP	771	0.1904	-7471A	PPB	25 May 2017 13:53:38	U	SMPL
I3331-06 - 1	SB-21COMP	290	0.0775	-7471A	PPB	25 May 2017 13:55:56	U	SMPL
I3331-06DUP - 1	SB-21COMPDUP	287	0.0768	-7471A	PPB	25 May 2017 13:58:14	U	SMPL
I3331-06MS - 1	SB-21COMPMS	15127	3.5606	-7471A	PPB	25 May 2017 14:00:30	U	SMPL
CCV20 - 1	CCV20	21556	5.0699	-7471A	PPB	25 May 2017 14:02:46	U	SMPL
CCB20 - 1	CCB20	48	0.0207	-7471A	PPB	25 May 2017 14:05:03	U	SMPL
I3331-06MSD - 1	SB-21COMPMSD	15236	3.5862	-7471A	PPB	25 May 2017 14:07:19	U	SMPL
I3331-10 - 1	SB-26COMP	262	0.0709	-7471A	PPB	25 May 2017 14:09:35	U	SMPL
I3332-02 - 1	SB-25COMP	846	0.2080	-7471A	PPB	25 May 2017 14:11:52	U	SMPL
I3332-06 - 1	SB-28COMP	341	0.0895	-7471A	PPB	25 May 2017 14:14:08	U	SMPL
I3332-08 - 1	SB-15COMP	205	0.0575	-7471A	PPB	25 May 2017 14:16:25	U	SMPL
I3339-01 - 1	E2-Q3-WSW	101967	23.9470	-7471A	PPB	25 May 2017 14:18:42	U	SMPL
I3339-02 - 1	E2-Q3-BASE	70708	16.6087	-7471A	PPB	25 May 2017 14:20:59	U	SMPL
I3339-03 - 1	E2-S10-BASE	19632	4.6182	-7471A	PPB	25 May 2017 14:23:18	U	SMPL
PB99287BL - 1	PBS	-27	0.0031	-7471A	PPB	25 May 2017 14:25:37	U	SMPL
PB99287BS - 1	LCSS	14849	3.4953	-7471A	PPB	25 May 2017 14:27:53	U	SMPL
CCV21 - 1	CCV21	20919	4.9203	-7471A	PPB	25 May 2017 14:30:09	U	SMPL
CCB21 - 1	CCB21	-157	-0.0275	-7471A	PPB	25 May 2017 14:32:26	U	SMPL
I3340-01 - 1	SP-1	1569	0.3777	-7471A	PPB	25 May 2017 14:34:42	U	SMPL
I3341-01 - 1	NM11-COMP-0-6	9702	2.2870	-7471A	PPB	25 May 2017 14:36:59	U	SMPL
I3341-02 - 1	NM11-COMP-6-18	14650	3.4486	-7471A	PPB	25 May 2017 14:39:15	U	SMPL
I3341-03 - 1	NM14-COMP-0-6	8221	1.9394	-7471A	PPB	25 May 2017 14:41:31	U	SMPL
I3341-04 - 1	NM14-COMP-6-18	7061	1.6670	-7471A	PPB	25 May 2017 14:43:48	U	SMPL
I3341-05 - 1	NM1-COMP-0-6	12316	2.9007	-7471A	PPB	25 May 2017 14:46:05	U	SMPL
I3341-06 - 1	NM1-COMP-6-18	6871	1.6224	-7471A	PPB	25 May 2017 14:48:22	U	SMPL
I3341-09 - 1	NM9-COMP-6-18	12122	2.8551	-7471A	PPB	25 May 2017 14:50:40	U	SMPL
I3341-10 - 1	NM9-COMP-0-6	10130	2.3875	-7471A	PPB	25 May 2017 14:52:58	U	SMPL
I3345-01 - 1	DUCT-BANK	1117	0.2716	-7471A	PPB	25 May 2017 14:55:14	U	SMPL
CCV22 - 1	CCV22	21329	5.0166	-7471A	PPB	25 May 2017 14:57:30	U	SMPL
CCB22 - 1	CCB22	36	0.0179	-7471A	PPB	25 May 2017 14:59:47	U	SMPL
I3345-01DUP - 1	DUCT-BANKDUP	1275	0.3087	-7471A	PPB	25 May 2017 15:02:03	U	SMPL
I3345-01MS - 1	DUCT-BANKMS	15832	3.7261	-7471A	PPB	25 May 2017 15:04:20	U	SMPL
I3345-01MSD - 1	DUCT-BANKMSD	16250	3.8242	-7471A	PPB	25 May 2017 15:06:36	U	SMPL
I3347-01 - 1	72-11931/VNJ-208/153/341/132	3054	0.7264	-7471A	PPB	25 May 2017 15:08:52	U	SMPL
I3347-05 - 1	284-CONCRETE	10224	2.4096	-7471A	PPB	25 May 2017 15:11:09	U	SMPL
I3348-01 - 1	VNJ-225/289/330/139/15532/VNJ-230	304573	71.5105	-7471A	PPB	25 May 2017 15:13:25	U	SMPL
I3349-01 - 1	TP-4	1474	0.3554	-7471A	PPB	25 May 2017 15:15:43	U	SMPL
PB99270BS - 1	LCSS	15230	3.5848	-7471A	PPB	25 May 2017 15:18:03	U	SMPL
I3331-06LX5 - 1		-16	0.0056	-7471A	PPB	25 May 2017 15:23:49	U	SMPL
I3331-06A - 1		16379	3.8545	-7471A	PPB	25 May 2017 15:26:04	U	SMPL
CCV23 - 1	CCV23	21194	4.9849	-7471A	PPB	25 May 2017 15:28:21	U	SMPL
CCB23 - 1	CCB23	-13	0.0063	-7471A	PPB	25 May 2017 15:30:37	U	SMPL
I3345-01LX5 - 1		234	0.0643	-7471A	PPB	25 May 2017 15:32:54	U	SMPL
I3345-01A - 1		17854	4.2008	-7471A	PPB	25 May 2017 15:35:11	U	SMPL
I3329-03DLX5 - 1		28570	6.7165	-7471A	PPB	25 May 2017 15:37:28	U	SMPL
I3329-08DLX10 - 1		29937	7.0374	-7471A	PPB	25 May 2017 15:39:44	U	SMPL
I3339-01DLX5 - 1		20034	4.7126	-7471A	PPB	25 May 2017 15:42:01	U	SMPL
I3339-02DLX5 - 1		13841	3.2587	-7471A	PPB	25 May 2017 15:44:18	U	SMPL
I3348-01DLX10 - 1		32737	7.6947	-7471A	PPB	25 May 2017 15:46:35	U	SMPL
CCV24 - 1	CCV24	21413	5.0363	-7471A	PPB	25 May 2017 15:50:44	U	SMPL
CCB24 - 1	CCB24	19	0.0139	-7471A	PPB	25 May 2017 15:55:36	U	SMPL