



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87699

Temp.: 19.6 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/25/2017	12:05
2	CAL2	1	Water			7.01	05/25/2017	12:06
3	CAL3	1	Water			10.02	05/25/2017	12:07
4	ICV	1	Water			6.99	05/25/2017	12:09
5	CCV1	1	Water			2.02	05/25/2017	12:10
6	I3340-09	1	Water			7.53	05/25/2017	12:11
7	I3340-09DUP	1	Water			7.54	05/25/2017	12:12
8	I3340-13	1	Water			6.99	05/25/2017	12:14
9	CCV2	1	Water			12.01	05/25/2017	12:15

23 87699

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3340

WorkList ID : 98912

Date : 5/25/2017 11:28:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/24/2017	Water	I3340-09-17	pH	Cool 4 deg C	PSEG01	H41	265563	05/24/2017	9045C
05/24/2017	Water	I3340-13-17	pH	Cool 4 deg C	PSEG01	H41	DRUM-1-DRUM-2-COMP	05/24/2017	9045C

Date/Time 03-23-17 11:28 AM
Received by: JP
Relinquished by: JP

Date/Time 05-23-17 hnm
Received by: JP
Relinquished by: JP