



Analytical Summary Report

Analysis Method: 9045C
Parameter: Corrosivity
Run Number: LB87701
BalanceID: WC SC-4

Analyst By : JIGNESH
Supervisor Review By : apatel
Temp.: 21.6 °C
SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/25/2017	12:05
2	CAL2	1	Water			7.01	05/25/2017	12:06
3	CAL3	1	Water			10.02	05/25/2017	12:07
4	ICV	1	Water			6.99	05/25/2017	12:09
5	CCV1	1	Water			2.02	05/25/2017	12:25
6	I3340-07	1	Solid	20.02	20	7.73	05/25/2017	12:26
7	I3340-07DUP	1	Solid	20.03	20	7.74	05/25/2017	12:27
8	I3345-07	1	Solid	20.04	20	7.86	05/25/2017	12:30
9	I3346-13	1	Solid	20.02	20	7.61	05/25/2017	12:31
10	I3347-04	1	Solid	20.01	20	9.14	05/25/2017	12:32
11	I3347-06	1	Solid	20.03	20	9.37	05/25/2017	12:33
12	I3348-03	1	Solid	20.04	20	8.50	05/25/2017	12:35
13	I3349-02	1	Solid	20.01	20	9.60	05/25/2017	12:36
14	CCV2	1	Water			12.01	05/25/2017	12:38

2B87701

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3340

WorkList ID : 98910

Date : 5/25/2017 11:27:12 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/31/2017	Solid	I3340-07	Corrosivity	Cool 4 deg C	PSEG01	H41	SP-1	05/24/2017	9045C
05/29/2017	Solid	I3345-07	Corrosivity	Cool 4 deg C	PSEG01	H42	DUCT-BANK	05/24/2017	9045C
05/31/2017	Solid	I3346-13	Corrosivity	Cool 4 deg C	PSEG02	H51	DRUM-7-8-11-13	05/24/2017	9045C
05/31/2017	Solid	I3347-04	Corrosivity	Cool 4 deg C	PSEG01	G61	72-11931/VNJ-208/153/341/13	05/24/2017	9045C
05/31/2017	Solid	I3347-06	Corrosivity	Cool 4 deg C	PSEG01	G61	284-CONCRETE	05/24/2017	9045C
05/31/2017	Solid	I3348-03	Corrosivity	Cool 4 deg C	PSEG01	H62	VNJ-225/289/330/139/15532A	05/24/2017	9045C
05/31/2017	Solid	I3349-02	Corrosivity	Cool 4 deg C	PSEG01	H43	TP-4	05/24/2017	9045C

05/25/17 11:30 AM

Date/Time
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 05.25.17 5:30 PM
Received by: [Signature]
Relinquished by: [Signature]