


245.1

LB87706
INSTRUMENT ID: CV2

Linear

 μ Abs.:

33015



A= 0.0000e+000

B= 3.0366e-004

C= -2.0155e-002

Rho= 0.9999399

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	95	0.000	95				
0.2	0.200	0.151	-0.049	565	0.0 %	565				
2.5	2.500	2.535	0.035	8414	0.0 %	8414				
5.0	5.000	5.055	0.055	16714	0.0 %	16714				
7.5	7.500	7.444	-0.056	24582	0.0 %	24582				
10.0	10.000	10.005	0.005	33015	0.0 %	33015				

LB87706 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 May 2017 16:10:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	95	-	0.0000	245.1	PPB	25 May 2017 16:11:56	S	Std
0.2 - 1	50.2	565	-	0.2000	245.1	PPB	25 May 2017 16:14:05	S	Std
2.5 - 1	52.5	8414	-	2.5000	245.1	PPB	25 May 2017 16:16:13	S	Std
5.0 - 1	55	16714	-	5.0000	245.1	PPB	25 May 2017 16:18:21	S	Std
7.5 - 1	57.5	24582	-	7.5000	245.1	PPB	25 May 2017 16:20:29	S	Std
10.0 - 1	510	33015	-	10.0000	245.1	PPB	25 May 2017 16:22:36	S	Std
ICV75 - 1	ICV75	12887	3.8931	-245.1		PPB	25 May 2017 16:25:19	U	SMPL
ICB75 - 1	ICB75	76	0.0029	-245.1		PPB	25 May 2017 16:27:25	U	SMPL
CCV25 - 1	CCV25	16610	5.0237	-245.1		PPB	25 May 2017 16:29:31	U	SMPL
CCB25 - 1	CCB25	60	-0.0019	-245.1		PPB	25 May 2017 16:31:38	U	SMPL
CRA - 1	CRA	579	0.1557	-245.1		PPB	25 May 2017 16:33:45	U	SMPL
HighStd - 1	HighStd	33414	10.1264	-245.1		PPB	25 May 2017 16:35:51	U	SMPL
ChkStd - 1	ChkStd	21852	6.6155	-245.1		PPB	25 May 2017 16:37:58	U	SMPL
PB99288BL - 1	PBW	3	-0.0192	-245.1		PPB	25 May 2017 16:40:06	U	SMPL
PB99288BS - 1	LCSW	14463	4.3717	-245.1		PPB	25 May 2017 16:42:13	U	SMPL
I3316-01 - 1	WS-3	-109	-0.0533	-245.1		PPB	25 May 2017 16:48:53	U	SMPL
I3316-01DUP - 1	WS-3DUP	-80	-0.0444	-245.1		PPB	25 May 2017 16:51:01	U	SMPL
I3316-01MS - 1	WS-3MS	15105	4.5667	-245.1		PPB	25 May 2017 16:53:10	U	SMPL
I3316-01MSD - 1	WS-3MSD	15653	4.7331	-245.1		PPB	25 May 2017 16:55:18	U	SMPL
CCV26 - 1	CCV26	16755	5.0677	-245.1		PPB	25 May 2017 16:57:25	U	SMPL
CCB26 - 1	CCB26	59	-0.0022	-245.1		PPB	25 May 2017 16:59:32	U	SMPL
I3316-01LX5 - 1		132	0.0199	-245.1		PPB	25 May 2017 17:01:39	U	SMPL
I3316-01A - 1		13486	4.0750	-245.1		PPB	25 May 2017 17:03:46	U	SMPL
CCV27 - 1	CCV27	16338	4.9411	-245.1		PPB	25 May 2017 17:05:52	U	SMPL
CCB27 - 1	CCB27	105	0.0117	-245.1		PPB	25 May 2017 17:07:59	U	SMPL