



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87712

Temp.: 20.6 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	05/25/2017	17:00
2	CAL2	1	Water			7.01	05/25/2017	17:01
3	CAL3	1	Water			10.03	05/25/2017	17:02
4	ICV	1	Water			7.01	05/25/2017	17:05
5	CCV1	1	Water			2.01	05/25/2017	17:06
6	I3371-02	1	Water			7.06	05/25/2017	17:09
7	I3371-02DUP	1	Water			7.07	05/25/2017	17:10
8	CCV2	1	Water			12.02	05/25/2017	17:11

LB 87712

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I3371 WorkList ID : 98931 Date : 5/25/2017 4:38:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/25/2017	Water	I3371-02-F	Corrosivity	Cool 4 deg C	PSEG01	H51	A598	05/25/2017	9040C

Date/Time 05/25/17 4:40 PM
Received by: JP
Relinquished by: CD

Date/Time 05/25/17 5:35 PM
Received by: CD
Relinquished by: SD