



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87714

Temp.: 20.6 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	05/25/2017	17:00
2	CAL2	1	Water			7.01	05/25/2017	17:01
3	CAL3	1	Water			10.03	05/25/2017	17:02
4	ICV	1	Water			7.01	05/25/2017	17:05
5	CCV1	1	Water			2.01	05/25/2017	17:06
6	I3366-01	1	Solid	20.02	20	5.89	05/25/2017	17:09
7	I3366-01DUP	1	Solid	20.04	20	5.90	05/25/2017	17:11
8	CCV2	1	Water			12.01	05/25/2017	17:12

1B 87714

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 13366

WorkList ID : 98930

Date : 5/25/2017 4:33:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3366-01-β	pH	Cool 4 deg C	TULL02	J61	FK-PS-01-004	05/25/2017	9045C

Date/Time 05/25/17 17:00M
Received by: JB
Relinquished by: [Signature]

Date/Time 05/25/17 15:35:00M
Received by: [Signature]
Relinquished by: JB