

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB87731

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP55256
calibration std. phosphate 1 ppm	WP55223
calibration std. phosphate 0.5 ppm	WP55222
calibration std. phosphate 0.3 ppm	WP55221
calibration std. 0 ppm	WP55218
phosphate ICV std.	WP55225
phosphate CCV std.	WP55224
calibration std. phosphate 0.1 ppm	WP55220
calibration std. phosphate 0.05 ppm	WP55219
Combined reagent	WP55227
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
Ascorbic Acid	WP55226

Intercept: -0.0009Slope: 0.6501Regression: 0.999809

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	05/26/2017	13:25
2	CAL2	0.05	1	50	50	0.032	05/26/2017	13:25
3	CAL3	0.10	1	50	50	0.065	05/26/2017	13:26
4	CAL4	0.30	1	50	50	0.186	05/26/2017	13:26
5	CAL5	0.50	1	50	50	0.331	05/26/2017	13:27
6	CAL6	1.00	1	50	50	0.648	05/26/2017	13:27



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB87731

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.494	05/26/2017	13:28
2	ICB		1	50	50	0.000	0.001	05/26/2017	13:28
3	CCV1	0.50	1	50	50	0.325	0.501	05/26/2017	13:29
4	CCB1		1	50	50	0.000	0.001	05/26/2017	13:29
5	PB99345BL		1	50	50	0.000	0.001	05/26/2017	13:30
6	PB99345BS	0.50	1	50	50	0.318	0.491	05/26/2017	13:30
7	I3326-01		1	50	50	2.228	3.429	05/26/2017	13:31
8	I3351-01		1	50	50	0.021	0.034	05/26/2017	13:31
9	I3351-01DUP		1	50	50	0.020	0.032	05/26/2017	13:32
10	I3351-01MS	0.50	1	50	50	0.351	0.541	05/26/2017	13:32
11	I3351-01MSD	0.50	1	50	50	0.345	0.532	05/26/2017	13:33
12	I3351-02		1	50	50	0.022	0.035	05/26/2017	13:33
13	I3326-01		100	50	50	0.426	0.657	05/26/2017	13:34
14	CCV2	0.50	1	50	50	0.326	0.503	05/26/2017	13:34
15	CCB2		1	50	50	0.000	0.001	05/26/2017	13:35