



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87736

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/26/2017	15:35
2	CAL2	1	Water			7.00	05/26/2017	15:36
3	CAL3	1	Water			10.02	05/26/2017	15:39
4	ICV	1	Water			6.99	05/26/2017	15:40
5	CCV1	1	Water			2.01	05/26/2017	15:41
6	I3365-03	1	Solid	20.02	20	8.95	05/26/2017	15:42
7	I3388-05	1	Solid	20.03	20	8.50	05/26/2017	15:43
8	I3388-06	1	Solid	20.02	20	8.48	05/26/2017	15:46
9	I3388-06DUP	1	Solid	20.01	20	8.49	05/26/2017	15:47
10	CCV2	1	Water			12.02	05/26/2017	15:49

1B8FF36

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3388			WorkList ID : 98970		Date : 5/26/2017 3:25:51 PM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/01/2017	Solid	I3365-03	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-03-COMP	05/25/2017	9045C
	Solid	I3388-05	Corrosivity	Cool 4 deg C	ENTA04		WC-B43-B50-002	05/26/2017	9045C
	Solid	I3388-06	Corrosivity	Cool 4 deg C	ENTA04		WC-B29-B35-002	05/26/2017	9045C

Date/Time 05/26/17 3:25 PM
Received by: JGR
Relinquished by:

Date/Time 05/26/17 5:30 PM
Received by: TR
Relinquished by: JP