


7470A

LB 87747
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35900



A= 0.0000e+000

B= 2.7968e-004

C= 4.9335e-002

Rho= 0.9998603

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	-12	0.000	-12				
0.2	0.200	0.239	0.039	679	0.0 %	679				
2.5	2.500	2.463	-0.037	8629	0.0 %	8629				
5.0	5.000	4.921	-0.079	17419	0.0 %	17419				
7.5	7.500	7.441	-0.059	26429	0.0 %	26429				
10.0	10.000	10.090	0.090	35900	0.0 %	35900				

LB87747 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 May 2017 09:03:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-12	-	0.0000	7470A	PPB	30 May 2017 09:30:35	S	Std
0.2 - 1	50.2	679	-	0.2000	7470A	PPB	30 May 2017 09:32:43	S	Std
2.5 - 1	52.5	8629	-	2.5000	7470A	PPB	30 May 2017 09:34:52	S	Std
5.0 - 1	55	17419	-	5.0000	7470A	PPB	30 May 2017 09:37:00	S	Std
7.5 - 1	57.5	26429	-	7.5000	7470A	PPB	30 May 2017 09:39:08	S	Std
10.0 - 1	510	35900	-	10.0000	7470A	PPB	30 May 2017 09:43:43	S	Std
ICV76 - 1	ICV76	14256	4.0365	-	7470A	PPB	30 May 2017 09:46:38	U	SMPL
ICB76 - 1	ICB76	-111	0.0183	-	7470A	PPB	30 May 2017 09:48:45	U	SMPL
CCV28 - 1	CCV28	17615	4.9759	-	7470A	PPB	30 May 2017 09:50:51	U	SMPL
CCB28 - 1	CCB28	-105	0.0200	-	7470A	PPB	30 May 2017 09:52:57	U	SMPL
CRA - 1	CRA	672	0.2373	-	7470A	PPB	30 May 2017 09:55:04	U	SMPL
HighStd - 1	HighStd	36170	10.1654	-	7470A	PPB	30 May 2017 09:57:10	U	SMPL
ChkStd - 1	ChkStd	26461	7.4500	-	7470A	PPB	30 May 2017 09:59:17	U	SMPL
PB99303BL - 1	PBW	-73	0.0289	-	7470A	PPB	30 May 2017 10:01:25	U	SMPL
PB99303BS - 1	LCSW	13698	3.8804	-	7470A	PPB	30 May 2017 10:03:32	U	SMPL
PB99303TB - 1	PB99260TB	-34	0.0398	-	7470A	PPB	30 May 2017 10:05:40	U	SMPL
I3331-16 - 1	SB-22COMP	-1	0.0491	-	7470A	PPB	30 May 2017 10:07:48	U	SMPL
I3331-16DUP - 1	SB-22COMP DUP	11	0.0524	-	7470A	PPB	30 May 2017 10:09:57	U	SMPL
I3331-16MS - 1	SB-22COMPMS	13425	3.8041	-	7470A	PPB	30 May 2017 10:12:05	U	SMPL
I3331-16MSD - 1	SB-22COMPMSD	12364	3.5073	-	7470A	PPB	30 May 2017 10:14:12	U	SMPL
CCV29 - 1	CCV29	17666	4.9902	-	7470A	PPB	30 May 2017 10:16:19	U	SMPL
CCB29 - 1	CCB29	-114	0.0175	-	7470A	PPB	30 May 2017 10:18:26	U	SMPL
I3331-17 - 1	SB-23COMP	0	0.0493	-	7470A	PPB	30 May 2017 10:20:32	U	SMPL
I3331-18 - 1	SB-21COMP	9	0.0519	-	7470A	PPB	30 May 2017 10:22:39	U	SMPL
I3331-19 - 1	SB-24COMP	15	0.0535	-	7470A	PPB	30 May 2017 10:24:45	U	SMPL
I3331-20 - 1	SB-26COMP	28	0.0572	-	7470A	PPB	30 May 2017 10:26:52	U	SMPL
I3332-13 - 1	SB-25COMP	20	0.0549	-	7470A	PPB	30 May 2017 10:28:59	U	SMPL
I3332-14 - 1	SB-27COMP	7	0.0513	-	7470A	PPB	30 May 2017 10:31:07	U	SMPL
I3332-15 - 1	SB-28COMP	4	0.0505	-	7470A	PPB	30 May 2017 10:33:15	U	SMPL
I3332-16 - 1	SB-15COMP	16	0.0538	-	7470A	PPB	30 May 2017 10:35:24	U	SMPL
I3331-16LX5 - 1		24	0.0560	-	7470A	PPB	30 May 2017 10:37:32	U	SMPL
I3331-16A - 1		14290	4.0460	-	7470A	PPB	30 May 2017 10:39:39	U	SMPL
CCV30 - 1	CCV30	17589	4.9687	-	7470A	PPB	30 May 2017 10:41:46	U	SMPL
CCB30 - 1	CCB30	-112	0.0180	-	7470A	PPB	30 May 2017 10:43:53	U	SMPL