


7471A

LB87749
INSTRUMENT ID: CY1

Linear

 μ Abs.:

41791



A= 0.0000e+000

B= 2.4086e-004

C= -1.3999e-002

Rho= 0.9999100

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	109	0.000	109				
0.2	0.200	0.187	-0.013	833	0.0 %	833				
2.5	2.500	2.510	0.010	10481	0.0 %	10481				
5.0	5.000	5.039	0.039	20979	0.0 %	20979				
7.5	7.500	7.400	-0.100	30780	0.0 %	30780				
10.0	10.000	10.052	0.052	41791	0.0 %	41791				

LB87749 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 30 May 2017 11:02:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	109	-	0.0000	7471A	PPB	30 May 2017 11:03:49	S	Std
0.2 - 1	50.2	833	-	0.2000	7471A	PPB	30 May 2017 11:06:07	S	Std
2.5 - 1	52.5	10481	-	2.5000	7471A	PPB	30 May 2017 11:08:25	S	Std
5.0 - 1	55	20979	-	5.0000	7471A	PPB	30 May 2017 11:10:42	S	Std
7.5 - 1	57.5	30780	-	7.5000	7471A	PPB	30 May 2017 11:12:59	S	Std
10.0 - 1	511	41791	-	10.0000	7471A	PPB	30 May 2017 11:17:53	S	Std
ICV77 - 1	ICV77	16514	3.9636	-	7471A	PPB	30 May 2017 11:20:31	U	SMPL
ICB77 - 1	ICB77	40	-0.0044	-	7471A	PPB	30 May 2017 11:22:47	U	SMPL
CCV31 - 1	CCV31	20911	5.0227	-	7471A	PPB	30 May 2017 11:25:02	U	SMPL
CCB31 - 1	CCB31	34	-0.0058	-	7471A	PPB	30 May 2017 11:27:18	U	SMPL
CRA - 1	CRA	813	0.1818	-	7471A	PPB	30 May 2017 11:29:34	U	SMPL
HighStd - 1	HighStd	41508	9.9837	-	7471A	PPB	30 May 2017 11:31:50	U	SMPL
ChkStd - 1	ChkStd	29334	7.0514	-	7471A	PPB	30 May 2017 11:34:06	U	SMPL
PB99309BL - 1	PBS	40	-0.0044	-	7471A	PPB	30 May 2017 11:36:22	U	SMPL
PB99309BS - 1	LCSS	14716	3.5305	-	7471A	PPB	30 May 2017 11:38:39	U	SMPL
I3366-01 - 1	FK-PS-01-004	996	0.2259	-	7471A	PPB	30 May 2017 11:40:56	U	SMPL
I3366-01DUP - 1	FK-PS-01-004DUP	994	0.2254	-	7471A	PPB	30 May 2017 11:45:49	U	SMPL
I3366-01MS - 1	FK-PS-01-004MS	15107	3.6247	-	7471A	PPB	30 May 2017 11:48:06	U	SMPL
I3366-01MSD - 1	FK-PS-01-004MSD	15657	3.7572	-	7471A	PPB	30 May 2017 11:50:24	U	SMPL
CCV32 - 1	CCV32	20852	5.0085	-	7471A	PPB	30 May 2017 11:52:40	U	SMPL
CCB32 - 1	CCB32	16	-0.0101	-	7471A	PPB	30 May 2017 11:54:56	U	SMPL
I3366-01LX5 - 1		261	0.0489	-	7471A	PPB	30 May 2017 11:57:12	U	SMPL
I3366-01A - 1		16818	4.0368	-	7471A	PPB	30 May 2017 12:02:07	U	SMPL
CCV33 - 1	CCV33	20658	4.9617	-	7471A	PPB	30 May 2017 12:04:22	U	SMPL
CCB33 - 1	CCB33	8	-0.0121	-	7471A	PPB	30 May 2017 12:06:38	U	SMPL