



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB87750

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	05/30/2017	09:10
2	I3316-01		1	>100.0	>212.0	05/30/2017	09:46
3	I3316-01DUP		1	>100.0	>212.0	05/30/2017	10:15
4	I3340-11		1	>100.0	>212.0	05/30/2017	10:45
5	I3340-16		1	>100.0	>212.0	05/30/2017	11:16
6	I3371-02		1	>100.0	>212.0	05/30/2017	11:45

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 87750

Worklist Name : FP53017

Worklist ID : 99016

Date : 5/30/2017 08:30AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3316-01	Flash Point	Cool 4 deg C	TULL02	I32	WS-3	05/23/2017	1010A
05/31/2017	Water	I3340-11 C	Flash Point	Cool 4 deg C	PSEG01	H41	265563	05/24/2017	1010A
05/31/2017	Water	I3340-16 C	Flash Point	Cool 4 deg C	PSEG01	H41	DRUM-1-DRUM-2-COMP	05/24/2017	1010A
05/30/2017	Water	I3371-02	Flash Point	Cool 4 deg C	PSEG01	H51	A598	05/25/2017	1010A

Date/Time 05/30/17 08:35 AM
 Received by: KS
 Relinquished by: CP

Date/Time 05/30/17 11:50 AM
 Received by: CP
 Relinquished by: KS