

0387753

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AD

5/30/2017 11:19

Rec'd
Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.600	0.0	0.096	
ICB1	-0.010	0.0	0.004	
CCV1	246.444	0.0	0.235	
CCB1	0.231	0.0	0.004	
PB99324BL	0.085	0.0	0.004	
I3348-03	-0.183	0.0	0.004	
I3348-03DUP	-0.508	0.0	0.004	
I3365-03	-0.311	0.0	0.004	
PB99325BL	-0.477	0.0	0.004	
I3371-02	-0.361	0.0	0.004	
I3371-02DUP	-0.337	0.0	0.004	
CCV2	259.425	0.0	0.247	
CCB2	0.032	0.0	0.004	

N	13
Mean	46.356
SD	95.6658
CV%	206.37

LIB 87 253

101

5/30/17

Aquakem v. 7.2AQ1

Results from time period:

Tue May 30 11:11:11 2017

Tue May 30 11:14:35 2017

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.0419	µg/l	5/30/2017 9:36:33	
5.0PPBCN	A	Reactive CN	P	4.7291	µg/l	5/30/2017 9:36:34	
10PPBCN	A	Reactive CN	P	9.7615	µg/l	5/30/2017 9:36:35	
50PPBCN	A	Reactive CN	P	50.5565	µg/l	5/30/2017 9:36:36	
100PPBCN	A	Reactive CN	P	100.4249	µg/l	5/30/2017 9:36:37	
250PPBCN	A	Reactive CN	P	249.406	µg/l	5/30/2017 9:36:38	
500PPBCN	A	Reactive CN	P	500.1638	µg/l	5/30/2017 9:36:39	
ICV1	S	Reactive CN	P	98.6005	µg/l	5/30/2017 11:11:11	
ICB1	S	Reactive CN	P	-0.01	µg/l	5/30/2017 11:11:12	
CCV1	S	Reactive CN	P	246.4443	µg/l	5/30/2017 11:11:13	
CCB1	S	Reactive CN	P	0.231	µg/l	5/30/2017 11:11:14	
PB99324BL	S	Reactive CN	P	0.0849	µg/l	5/30/2017 11:11:15	
I3348-03	S	Reactive CN	P	-0.1825	µg/l	5/30/2017 11:11:16	
I3348-03DI	S	Reactive CN	P	-0.5077	µg/l	5/30/2017 11:11:17	
I3365-03	S	Reactive CN	P	-0.311	µg/l	5/30/2017 11:11:18	
PB99325BL	S	Reactive CN	P	-0.477	µg/l	5/30/2017 11:11:19	
I3371-02	S	Reactive CN	P	-0.3607	µg/l	5/30/2017 11:11:20	
I3371-02DI	S	Reactive CN	P	-0.3372	µg/l	5/30/2017 11:11:21	
CCV2	S	Reactive CN	P	259.4252	µg/l	5/30/2017 11:14:34	
CCB2	S	Reactive CN	P	0.0325	µg/l	5/30/2017 11:14:35	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : 

5/30/2017 9:44


Test Total CN

Accepted 5/30/2017 9:44

Factor 1070
Bias 0.004

Coeff. of det. 0.999995

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.0419	0.0000	
2	5.0PPBCN	0.009	4.7291	5.0000	
3	10PPBCN	0.013	9.7615	10.0000	
4	50PPBCN	0.052	50.5565	50.0000	
5	100PPBCN	0.098	100.4249	100.0000	
6	250PPBCN	0.237	249.4060	250.0000	
7	500PPBCN	0.472	500.1638	500.0000	