



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87769

Temp.: 19.6 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/31/2017	11:25
2	CAL2	1	Water			7.01	05/31/2017	11:26
3	CAL3	1	Water			10.04	05/31/2017	11:28
4	ICV	1	Water			7.01	05/31/2017	11:30
5	CCV1	1	Water			2.01	05/31/2017	11:31
6	I3245-06	1	Water			4.50	05/31/2017	11:32
7	I3245-06DUP	1	Water			4.51	05/31/2017	11:33
8	CCV2	1	Water			12.01	05/31/2017	11:35

1B87769

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i3245-1 WorkList ID : 99031 Date : 5/31/2017 11:06:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/18/2017	Water	I3245-06-F	pH	Cool 4 deg C	WOOD06	I61	FB	05/18/2017	SM 4500-PH B

Date/Time 05.31.17 11:06 AM
Received by: JP
Relinquished by:

Date/Time 05.31.17 2:00 PM
Received by: [Signature]
Relinquished by: [Signature]