



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87770

Temp.: 21.6 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/31/2017	11:25
2	CAL2	1	Water			7.01	05/31/2017	11:26
3	CAL3	1	Water			10.04	05/31/2017	11:28
4	ICV	1	Water			7.01	05/31/2017	11:30
5	CCV1	1	Water			2.02	05/31/2017	11:40
6	I3398-09	1	Solid	20.02	20	7.92	05/31/2017	11:41
7	I3398-09DUP	1	Solid	20.03	20	7.93	05/31/2017	11:42
8	I3398-10	1	Solid	20.04	20	7.47	05/31/2017	11:44
9	I3398-11	1	Solid	20.02	20	7.09	05/31/2017	11:45
10	I3398-12	1	Solid	20.03	20	8.33	05/31/2017	11:46
11	CCV2	1	Water			12.02	05/31/2017	11:48

LB87770

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3399 WorkList ID : 99032 Date : 5/31/2017 11:06:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3398-09	Corrosivity	Cool 4 deg C	LIRO01	J22	SB-08-COMP	05/26/2017	9045C
	Solid	I3398-10	Corrosivity	Cool 4 deg C	LIRO01	J22	SB-02-COMP	05/30/2017	9045C
	Solid	I3398-11	Corrosivity	Cool 4 deg C	LIRO01	J22	SB-01-COMP	05/30/2017	9045C
	Solid	I3398-12	Corrosivity	Cool 4 deg C	LIRO01	J22	SB-05-COMP	05/30/2017	9045C

Date/Time 05-31-17 11:06 AM
Received by: JP
Relinquished by: JP

Date/Time 05-31-17 3:30 PM
Received by: JP
Relinquished by: JP