


ISM02.4

LB87774
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43547



A= 0.0000e+000

B= 2.3043e-004

C= 1.6460e-002

Rho= 0.9999300

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.026	0.026	43	0.000	43					1
0.2	0.200	0.225	0.025	906	0.0 %	906					13
2.5	2.500	2.494	-0.006	10753	0.0 %	10753					0
5.0	5.000	4.914	-0.086	21255	0.0 %	21255					-2
7.5	7.500	7.489	-0.011	32429	0.0 %	32429					0
10.0	10.000	10.051	0.051	43547	0.0 %	43547					1

MB
5/3/12

LB87774
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	43	-		0 ISM02.4	PPB	5/31/2017 13:56	Std
	0.2 S0.2	906	-		0.2 ISM02.4	PPB	5/31/2017 13:58	Std
	2.5 S2.5	10753	-		2.5 ISM02.4	PPB	5/31/2017 14:01	Std
	5 S5	21255	-		5 ISM02.4	PPB	5/31/2017 14:03	Std
	7.5 S7.5	32429	-		7.5 ISM02.4	PPB	5/31/2017 14:05	Std
	10 S10	43547	-		10 ISM02.4	PPB	5/31/2017 14:07	Std
ICV	ICV	16857	3.9008	-	ISM02.4	PPB	5/31/2017 14:10	SMPL
ICB	ICB	-32	0.0091	-	ISM02.4	PPB	5/31/2017 14:12	SMPL
CCV043	CCV043	21270	4.9177	-	ISM02.4	PPB	5/31/2017 14:15	SMPL
CCB043	CCB043	-99	-0.0064	-	ISM02.4	PPB	5/31/2017 14:17	SMPL
PB99386BL	PBW003	-34	0.0086	-	ISM02.4	PPB	5/31/2017 14:19	SMPL
I3358-01	MCOL96	55	0.0291	-	ISM02.4	PPB	5/31/2017 14:21	SMPL
I3358-02	MCOLA0	40	0.0257	-	ISM02.4	PPB	5/31/2017 14:24	SMPL
I3358-03	MCOLB8	43	0.0264	-	ISM02.4	PPB	5/31/2017 14:26	SMPL
I3358-04	MCOLC2	41	0.0259	-	ISM02.4	PPB	5/31/2017 14:28	SMPL
I3358-05	MCOLC2D	43	0.0264	-	ISM02.4	PPB	5/31/2017 14:31	SMPL
I3358-06	MCOLC2S	4240	0.9935	-	ISM02.4	PPB	5/31/2017 14:33	SMPL
I3358-07	MCOLE4	-8	0.0146	-	ISM02.4	PPB	5/31/2017 14:35	SMPL
I3358-08	MCOLD4	28	0.0229	-	ISM02.4	PPB	5/31/2017 14:37	SMPL
I3358-09	MCOLD6	328	0.092	-	ISM02.4	PPB	5/31/2017 14:40	SMPL
I3358-10	MCOLE0	20	0.0211	-	ISM02.4	PPB	5/31/2017 14:42	SMPL
I3358-11	MCOLE2	-35	0.0084	-	ISM02.4	PPB	5/31/2017 14:44	SMPL
I3358-12	MCOLE9	-58	0.0031	-	ISM02.4	PPB	5/31/2017 14:46	SMPL
I3358-13	MCOLF0	-10	0.0142	-	ISM02.4	PPB	5/31/2017 14:49	SMPL
I3358-14	MCOLF3	65	0.0314	-	ISM02.4	PPB	5/31/2017 14:51	SMPL
I3358-15	MCOLF4	77	0.0342	-	ISM02.4	PPB	5/31/2017 14:53	SMPL
I3358-16	MCOLF5	87	0.0365	-	ISM02.4	PPB	5/31/2017 14:56	SMPL
I3358-17	MCOLF6	102	0.04	-	ISM02.4	PPB	5/31/2017 14:58	SMPL
I3358-18	MCOLF9	107	0.0411	-	ISM02.4	PPB	5/31/2017 15:00	SMPL
I3358-19	MCOLG0	81	0.0351	-	ISM02.4	PPB	5/31/2017 15:02	SMPL
CCV044	CCV044	20804	4.8103	-	ISM02.4	PPB	5/31/2017 15:05	SMPL
CCB044	CCB044	-87	-0.0036	-	ISM02.4	PPB	5/31/2017 15:07	SMPL
I3358-20	MCOLG1	-41	0.007	-	ISM02.4	PPB	5/31/2017 15:09	SMPL
I3358-21	MCOLG4	43	0.0264	-	ISM02.4	PPB	5/31/2017 15:12	SMPL
I3358-22	MCOLD2	48	0.0275	-	ISM02.4	PPB	5/31/2017 15:14	SMPL
PB99387BL	PBW003	66	0.0317	-	ISM02.4	PPB	5/31/2017 15:16	SMPL
I3359-01	MCOLB0	44	0.0266	-	ISM02.4	PPB	5/31/2017 15:18	SMPL
I3359-02	MCOLB4	30	0.0234	-	ISM02.4	PPB	5/31/2017 15:21	SMPL
I3359-03	MCOLB6	-80	-0.002	-	ISM02.4	PPB	5/31/2017 15:23	SMPL
I3359-04	MCOLB6D	-61	0.0024	-	ISM02.4	PPB	5/31/2017 15:25	SMPL
I3359-05	MCOLB6S	4272	1.0008	-	ISM02.4	PPB	5/31/2017 15:27	SMPL
I3359-06	MCOLC8	40	0.0257	-	ISM02.4	PPB	5/31/2017 15:30	SMPL
I3359-07	MCOLA2	70	0.0326	-	ISM02.4	PPB	5/31/2017 15:32	SMPL
I3359-08	MCOLC0	48	0.0275	-	ISM02.4	PPB	5/31/2017 15:34	SMPL
I3359-09	MCOLC4	22	0.0215	-	ISM02.4	PPB	5/31/2017 15:37	SMPL
I3359-10	MCOLC6	-65	0.0015	-	ISM02.4	PPB	5/31/2017 15:39	SMPL

I3359-11	MCOL94	-70	0.0003 -	ISM02.4	PPB	5/31/2017 15:41	SMPL
I3359-12	MCOL98	5	0.0176 -	ISM02.4	PPB	5/31/2017 15:43	SMPL
I3359-13	MCOLA4	29	0.0231 -	ISM02.4	PPB	5/31/2017 15:46	SMPL
I3359-14	MCOLA8	76	0.034 -	ISM02.4	PPB	5/31/2017 15:48	SMPL
I3359-15	MCOLB2	49	0.0278 -	ISM02.4	PPB	5/31/2017 15:50	SMPL
CCV045	CCV045	21028	4.8619 -	ISM02.4	PPB	5/31/2017 15:53	SMPL
CCB045	CCB045	-13	0.0135 -	ISM02.4	PPB	5/31/2017 15:55	SMPL
I3359-16	MCOLD0	-49	0.0052 -	ISM02.4	PPB	5/31/2017 15:57	SMPL
I3359-17	MCOLE6	-67	0.001 -	ISM02.4	PPB	5/31/2017 15:59	SMPL
I3359-18	MCOLG8	-12	0.0137 -	ISM02.4	PPB	5/31/2017 16:02	SMPL
I3359-19	MCOLA6	74	0.0335 -	ISM02.4	PPB	5/31/2017 16:04	SMPL
I3359-20	MCOLG6	59	0.0301 -	ISM02.4	PPB	5/31/2017 16:06	SMPL
PB99388BL	PBW003	64	0.0312 -	ISM02.4	PPB	5/31/2017 16:09	SMPL
I3360-01	MCOL95	-33	0.0089 -	ISM02.4	PPB	5/31/2017 16:11	SMPL
I3360-02	MCOL99	-67	0.001 -	ISM02.4	PPB	5/31/2017 16:13	SMPL
I3360-03	MCOLB7	-70	0.0003 -	ISM02.4	PPB	5/31/2017 16:15	SMPL
I3360-04	MCOLC1	34	0.0243 -	ISM02.4	PPB	5/31/2017 16:18	SMPL
I3360-05	MCOLE3	55	0.0291 -	ISM02.4	PPB	5/31/2017 16:20	SMPL
I3360-06	MCOLD3	32	0.0238 -	ISM02.4	PPB	5/31/2017 16:22	SMPL
I3360-07	MCOLD5	27	0.0227 -	ISM02.4	PPB	5/31/2017 16:24	SMPL
I3360-08	MCOLD7	-70	0.0003 -	ISM02.4	PPB	5/31/2017 16:27	SMPL
I3360-09	MCOLD8	15	0.0199 -	ISM02.4	PPB	5/31/2017 16:29	SMPL
I3360-10	MCOLD9	-59	0.0029 -	ISM02.4	PPB	5/31/2017 16:31	SMPL
I3360-11	MCOLE1	22	0.0215 -	ISM02.4	PPB	5/31/2017 16:34	SMPL
I3360-12	MCOLA9	40	0.0257 -	ISM02.4	PPB	5/31/2017 16:36	SMPL
I3360-13	MCOLB3	38	0.0252 -	ISM02.4	PPB	5/31/2017 16:38	SMPL
I3360-14	MCOLB5	21	0.0213 -	ISM02.4	PPB	5/31/2017 16:40	SMPL
CCV046	CCV046	21020	4.8601 -	ISM02.4	PPB	5/31/2017 16:43	SMPL
CCB046	CCB046	-65	0.0015 -	ISM02.4	PPB	5/31/2017 16:45	SMPL
I3360-15	MCOLC7	-57	0.0033 -	ISM02.4	PPB	5/31/2017 16:47	SMPL
I3360-16	MCOLD1	35	0.0245 -	ISM02.4	PPB	5/31/2017 16:50	SMPL
I3360-17	MCOLA1	39	0.0254 -	ISM02.4	PPB	5/31/2017 16:52	SMPL
I3360-18	MCOLA1D	40	0.0257 -	ISM02.4	PPB	5/31/2017 16:54	SMPL
I3360-19	MCOLA1S	4325	1.0131 -	ISM02.4	PPB	5/31/2017 16:56	SMPL
I3360-20	MCOLB9	-68	0.0008 -	ISM02.4	PPB	5/31/2017 16:59	SMPL
I3360-21	MCOLC3	-49	0.0052 -	ISM02.4	PPB	5/31/2017 17:01	SMPL
I3360-22	MCOLC5	-37	0.0079 -	ISM02.4	PPB	5/31/2017 17:03	SMPL
CCV047	CCV047	21074	4.8725 -	ISM02.4	PPB	5/31/2017 17:06	SMPL
CCB047	CCB047	10	0.0188 -	ISM02.4	PPB	5/31/2017 17:08	SMPL