

===== CB 8-17-95 =====
Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AV

6/1/2017 12:51

Test: Phenolics-

Sample Id	Result <i>wg/L</i>	Dil. 1 +	Response	Errors
ICV1	0.974	0.0	0.107	
ICB1	0.003	0.0	0.002	
CCV1	0.976	0.0	0.107	
CCB1	0.005	0.0	0.002	
PB99382BL	0.007	0.0	0.002	
PB99382BS	0.980	0.0	0.107	
I3309-01	0.009	0.0	0.003	
I3309-01DUP	0.007	0.0	0.002	
I3309-01MS	0.966	0.0	0.106	
I3309-01MSD	0.961	0.0	0.105	
I3309-02	0.009	0.0	0.003	
CCV2	0.974	0.0	0.107	
CCB2	0.002	0.0	0.002	

N	13
Mean	0.452
SD	0.5013
CV%	110.97

LB 87795

ALC

Aquakem v. 7.2AQ1

Results from time period:

Thu Jun 01 11:59:43 2017

Thu Jun 01 12:35:41 2017

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0533	mg/l	6/1/2017 11:59:43	
0.1PPM	A	Phenolics-	P	0.105	mg/l	6/1/2017 11:59:44	
0.25PPM	A	Phenolics-	P	0.2519	mg/l	6/1/2017 11:59:45	
0.50PPM	A	Phenolics-	P	0.4974	mg/l	6/1/2017 11:59:46	
1.0PPM	A	Phenolics-	P	0.9845	mg/l	6/1/2017 11:59:47	
2.0PPM	A	Phenolics-	P	2.0078	mg/l	6/1/2017 11:59:48	
ICV1	S	Phenolics-	P	0.9742	mg/l	6/1/2017 12:31:42	
ICB1	S	Phenolics-	P	0.0034	mg/l	6/1/2017 12:31:43	
CCV1	S	Phenolics-	P	0.9756	mg/l	6/1/2017 12:31:44	
CCB1	S	Phenolics-	P	0.0047	mg/l	6/1/2017 12:31:45	
PB99382BL	S	Phenolics-	P	0.0065	mg/l	6/1/2017 12:31:46	
PB99382BS	S	Phenolics-	P	0.9797	mg/l	6/1/2017 12:31:47	
I3309-01	S	Phenolics-	P	0.009	mg/l	6/1/2017 12:31:48	
I3309-01DUP	S	Phenolics-	P	0.007	mg/l	6/1/2017 12:31:49	
I3309-01MS	S	Phenolics-	P	0.9663	mg/l	6/1/2017 12:31:50	
I3309-01MSD	S	Phenolics-	P	0.9612	mg/l	6/1/2017 12:31:51	
I3309-02	S	Phenolics-	P	0.0088	mg/l	6/1/2017 12:31:52	
CCV2	S	Phenolics-	P	0.9744	mg/l	6/1/2017 12:31:53	
CCB2	S	Phenolics-	P	0.0016	mg/l	6/1/2017 12:35:40	
LCS	S	Phenolics-	P	0.9408	mg/l	6/1/2017 12:35:41	

LB87795

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : _____

6/1/2017 12:01


Test Phenolics-

Accepted 6/1/2017 12:01

Factor 9.272
Bias 0.002

Coeff. of det. 0.999876

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.007	0.0533	0.0500	
2	Phenol-2	0.013	0.1050	0.1000	
3	Phenol-2	0.029	0.2519	0.2500	
4	Phenol-2	0.055	0.4974	0.5000	
5	Phenol-2	0.108	0.9845	1.0000	
6	Phenol-2	0.218	2.0078	2.0000	