



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 06/02/2017DATE & TIME IN: 06/01/2017 16:35OVEN TEMP: 98 °CRun Number: LB87801DATE & TIME OUT: 06/02/2017 08:20OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB87801BL	LB87801BL	51.9631	51.9631	100	51.9631	51.9631	51.9631	0	0
2	I3414-01	TE-PMW-11S	54.0389	54.0389	100	54.0814	54.0814	54.0814	0.0425	425
3	I3414-01DUP	TE-PMW-11SDUP	56.7341	56.7341	100	56.7771	56.7771	56.7771	0.043	430

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

2387801

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS I3414

WorkList ID : 99094

Date : 6/1/2017 3:15:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3414-01-5	TS	Cool 4 deg C	TECT01	J51	TE-PMW-11S	05/31/2017	SM2540 B

Date/Time 06-01-17 3:15 PM
Received by: JP
Relinquished by: JP

Date/Time 06-01-17 3:00 PM
Received by: JP
Relinquished by: JP