



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87808

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/01/2017	16:25
2	CAL2	1	Water			7.01	06/01/2017	16:27
3	CAL3	1	Water			10.03	06/01/2017	16:28
4	ICV	1	Water			7.00	06/01/2017	16:30
5	CCV1	1	Water			2.01	06/01/2017	16:31
6	I3436-11	1	Solid	20.02	20	7.80	06/01/2017	16:32
7	I3436-11DUP	1	Solid	20.03	20	7.81	06/01/2017	16:33
8	I3436-12	1	Solid	20.04	20	6.90	06/01/2017	16:35
9	I3436-13	1	Solid	20.01	20	8.07	06/01/2017	16:36
10	I3436-14	1	Solid	20.02	20	7.97	06/01/2017	16:37
11	I3436-15	1	Solid	20.01	20	5.77	06/01/2017	16:39
12	I3438-05	1	Solid	20.04	20	6.16	06/01/2017	16:40
13	I3438-10	1	Solid	20.05	20	6.32	06/01/2017	16:41
14	I3438-15	1	Solid	20.01	20	7.51	06/01/2017	16:42
15	CCV2	1	Water			12.01	06/01/2017	16:43

WORKLIST(Hardcopy Internal Chain)

1987808

WorkList Name : CORROSIONITY I2436

WorkList ID : 99096

Date : 6/1/2017 3:55:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3436-11	Corrosivity	Cool 4 deg C	LIRO01	K22	SB-04-COMP	05/31/2017	9045C
	Solid	I3436-12	Corrosivity	Cool 4 deg C	LIRO01	K22	SB-06-COMP	05/31/2017	9045C
	Solid	I3436-13	Corrosivity	Cool 4 deg C	LIRO01	K22	SB-07-COMP	05/31/2017	9045C
	Solid	I3436-14	Corrosivity	Cool 4 deg C	LIRO01	K22	SB-09-COMP	05/31/2017	9045C
	Solid	I3436-15	Corrosivity	Cool 4 deg C	LIRO01	K22	SB-10-COMP	05/31/2017	9045C
06/08/2017	Solid	I3438-05	Corrosivity	Cool 4 deg C	PSEG01	J62	TP-5-6	06/01/2017	9045C
06/08/2017	Solid	I3438-10	Corrosivity	Cool 4 deg C	PSEG01	J62	TP-7	06/01/2017	9045C
06/08/2017	Solid	I3438-15	Corrosivity	Cool 4 deg C	PSEG01	J62	TP-8-9	06/01/2017	9045C

Date/Time

Received by:

Relinquished by:

06-01-17 3:55 PM

11

Date/Time

Received by:

Relinquished by:

06-01-17 6:35 PM

08

30