


7470A

LB87811
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35620



A= 0.0000e+000

B= 2.7975e-004

C= 5.5497e-003

Rho= 0.9999664

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	44	0.000	44				
0.2	0.200	0.190	-0.010	659	0.0 %	659				
2.5	2.500	2.456	-0.044	8758	0.0 %	8758				
5.0	5.000	5.037	0.037	17986	0.0 %	17986				
7.5	7.500	7.529	0.029	26893	0.0 %	26893				
10.0	10.000	9.970	-0.030	35620	0.0 %	35620				

LB87811 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 02 Jun 2017 09:50:49				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	44	-	0.00007470A	PPB	02 Jun 2017 10:18:32	S	Std
0.2 - 1	50.2	659	-	0.20007470A	PPB	02 Jun 2017 10:20:40	S	Std
2.5 - 1	52.5	8758	-	2.50007470A	PPB	02 Jun 2017 10:22:48	S	Std
5.0 - 1	55	17986	-	5.00007470A	PPB	02 Jun 2017 10:24:57	S	Std
7.5 - 1	57.5	26893	-	7.50007470A	PPB	02 Jun 2017 10:27:05	S	Std
10.0 - 1	50	35620	-	10.00007470A	PPB	02 Jun 2017 10:29:12	S	Std
ICV81 - 1	ICV81	14374	4.0267	-7470A	PPB	02 Jun 2017 10:32:29	U	SMPL
ICB81 - 1	ICB81	-25	-0.0014	-7470A	PPB	02 Jun 2017 10:34:35	U	SMPL
CCV52 - 1	CCV52	18014	5.0450	-7470A	PPB	02 Jun 2017 10:36:42	U	SMPL
CCB52 - 1	CCB52	-34	-0.0040	-7470A	PPB	02 Jun 2017 10:38:48	U	SMPL
CRA - 1	CRA	641	0.1849	-7470A	PPB	02 Jun 2017 10:40:55	U	SMPL
HighStd - 1	HighStd	35176	9.8462	-7470A	PPB	02 Jun 2017 10:43:01	U	SMPL
ChkStd - 1	ChkStd	25048	7.0128	-7470A	PPB	02 Jun 2017 10:45:08	U	SMPL
PB99407BL - 1	PBW	-38	-0.0051	-7470A	PPB	02 Jun 2017 10:47:16	U	SMPL
PB99407BS - 1	LCSW	14241	3.9895	-7470A	PPB	02 Jun 2017 10:49:24	U	SMPL
PB99407TB - 1	PB99394TB	-15	0.0014	-7470A	PPB	02 Jun 2017 10:51:31	U	SMPL
I3140-19 - 1	TP-3-3	27	0.0131	-7470A	PPB	02 Jun 2017 10:53:40	U	SMPL
I3140-19DUP - 1	TP-3-3DUP	27	0.0131	-7470A	PPB	02 Jun 2017 10:55:48	U	SMPL
I3140-19MS - 1	TP-3-3MS	13907	3.8961	-7470A	PPB	02 Jun 2017 10:57:56	U	SMPL
I3140-19MSD - 1	TP-3-3MSD	13293	3.7243	-7470A	PPB	02 Jun 2017 11:00:03	U	SMPL
CCV53 - 1	CCV53	18435	5.1628	-7470A	PPB	02 Jun 2017 11:02:10	U	SMPL
CCB53 - 1	CCB53	-32	-0.0034	-7470A	PPB	02 Jun 2017 11:04:17	U	SMPL
I3398-09 - 1	SB-08-COMP	25	0.0125	-7470A	PPB	02 Jun 2017 11:06:24	U	SMPL
I3398-10 - 1	SB-02-COMP	28	0.0134	-7470A	PPB	02 Jun 2017 11:08:30	U	SMPL
I3398-11 - 1	SB-01-COMP	40	0.0167	-7470A	PPB	02 Jun 2017 11:10:37	U	SMPL
I3398-12 - 1	SB-05-COMP	34	0.0151	-7470A	PPB	02 Jun 2017 11:12:44	U	SMPL
I3409-07 - 1	WC-NP-ESB-21-22-23-24	38	0.0162	-7470A	PPB	02 Jun 2017 11:14:52	U	SMPL
I3409-08 - 1	WC-NP-ESB-25-26-27	63	0.0232	-7470A	PPB	02 Jun 2017 11:16:59	U	SMPL
I3412-03 - 1	TP-1-TP-2-COMP	48	0.0190	-7470A	PPB	02 Jun 2017 11:19:07	U	SMPL
I3412-07 - 1	TP-3-TP-4-COMP	83	0.0288	-7470A	PPB	02 Jun 2017 11:21:16	U	SMPL
I3412-11 - 1	TP-5-COMP	86	0.0296	-7470A	PPB	02 Jun 2017 11:23:24	U	SMPL
I3413-03 - 1	R4-COMP	79	0.0277	-7470A	PPB	02 Jun 2017 11:25:31	U	SMPL
CCV54 - 1	CCV54	18206	5.0988	-7470A	PPB	02 Jun 2017 11:27:38	U	SMPL
CCB54 - 1	CCB54	-33	-0.0037	-7470A	PPB	02 Jun 2017 11:29:45	U	SMPL
I3420-05 - 1	TP-3-TP-4	41	0.0170	-7470A	PPB	02 Jun 2017 11:31:52	U	SMPL
I3420-10 - 1	TP-1-TP-2	35	0.0153	-7470A	PPB	02 Jun 2017 11:33:59	U	SMPL
I3140-19LX5 - 1		25	0.0125	-7470A	PPB	02 Jun 2017 11:36:06	U	SMPL
I3140-19A - 1		14068	3.9411	-7470A	PPB	02 Jun 2017 11:38:14	U	SMPL
CCV55 - 1	CCV55	18188	5.0937	-7470A	PPB	02 Jun 2017 11:40:21	U	SMPL
CCB55 - 1	CCB55	-39	-0.0054	-7470A	PPB	02 Jun 2017 11:42:29	U	SMPL