



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87824

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/02/2017	12:25
2	CAL2	1	Water			7.00	06/02/2017	12:26
3	CAL3	1	Water			10.03	06/02/2017	12:37
4	ICV	1	Water			7.01	06/02/2017	12:39
5	CCV1	1	Water			2.01	06/02/2017	12:40
6	I3437-06	1	Solid	20.02	20	11.35	06/02/2017	12:41
7	I3437-06DUP	1	Solid	20.03	20	11.37	06/02/2017	12:42
8	I3440-02	1	Solid	20.01	20	9.32	06/02/2017	12:45
9	I3440-04	1	Solid	20.03	20	9.34	06/02/2017	12:46
10	I3440-06	1	Solid	20.04	20	8.71	06/02/2017	12:47
11	I3441-02	1	Solid	20.01	20	9.02	06/02/2017	12:48
12	I3441-04	1	Solid	20.05	20	8.88	06/02/2017	12:49
13	I3441-06	1	Solid	20.02	20	8.67	06/02/2017	12:50
14	I3441-08	1	Solid	20.04	20	8.87	06/02/2017	12:51
15	I3441-10	1	Solid	20.01	20	8.56	06/02/2017	12:52
16	CCV2	1	Water			12.01	06/02/2017	12:53
17	I3441-12	1	Solid	20.02	20	7.71	06/02/2017	12:54
18	I3442-03	1	Solid	20.02	20	7.96	06/02/2017	12:55
19	I3443-02	1	Solid	20.03	20	5.58	06/02/2017	12:56
20	I3444-02	1	Solid	20.05	20	8.25	06/02/2017	12:57

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3444-04	1	Solid	20.05	20	7.61	06/02/2017	12:59
22	I3444-06	1	Solid	20.01	20	6.24	06/02/2017	13:00
23	I3444-08	1	Solid	20.02	20	7.23	06/02/2017	13:02
24	I3449-05	1	Solid	20.04	20	8.45	06/02/2017	13:03
25	I3449-05DUP	1	Solid	20.04	20	8.46	06/02/2017	13:05
26	CCV3	1	Water			12.02	06/02/2017	13:06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3440

WorkList ID : 99130

Date : 6/2/2017 11:49:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/06/2017	Solid	I3437-06	Corrosivity	Cool 4 deg C	TECT01	J51	COMP-1	06/01/2017	9045C
06/06/2017	Solid	I3440-02	Corrosivity	Cool 4 deg C	PSEG05	J11	EO-02-RO-4146-060117	06/01/2017	9045C
06/06/2017	Solid	I3440-04	Corrosivity	Cool 4 deg C	PSEG05	J11	EO-02-RO-5078-060117	06/01/2017	9045C
06/06/2017	Solid	I3440-06	Corrosivity	Cool 4 deg C	PSEG05	J11	EO-03-RO-6703-060117	06/01/2017	9045C
06/06/2017	Solid	I3441-02	Corrosivity	Cool 4 deg C	PSEG05	J12	OR-RO-5267-060117	06/01/2017	9045C
06/06/2017	Solid	I3441-04	Corrosivity	Cool 4 deg C	PSEG05	J12	OR-RO-5004-060117	06/01/2017	9045C
06/06/2017	Solid	I3441-06	Corrosivity	Cool 4 deg C	PSEG05	J12	OR-RO-141-060117	06/01/2017	9045C
06/06/2017	Solid	I3441-08	Corrosivity	Cool 4 deg C	PSEG05	J12	OR-RO-6710-060117	06/01/2017	9045C
06/06/2017	Solid	I3441-10	Corrosivity	Cool 4 deg C	PSEG05	J12	OR-RO-6789-060117	06/01/2017	9045C
06/06/2017	Solid	I3441-12	Corrosivity	Cool 4 deg C	PSEG05	J12	OR-RO-17273-060117	06/01/2017	9045C
06/06/2017	Solid	I3442-03	Corrosivity	Cool 4 deg C	PSEG05	J12	OK-02-060117	06/01/2017	9045C
06/06/2017	Solid	I3443-02	Corrosivity	Cool 4 deg C	PSEG05	J13	SA-RO-6699AA-060117	06/01/2017	9045C
06/06/2017	Solid	I3444-02	Corrosivity	Cool 4 deg C	PSEG05	J13	BU-RO-238-060117	06/01/2017	9045C
06/06/2017	Solid	I3444-04	Corrosivity	Cool 4 deg C	PSEG05	J13	BU-RO-240-060117	06/01/2017	9045C
06/06/2017	Solid	I3444-06	Corrosivity	Cool 4 deg C	PSEG05	J13	BU-RO-286-060117	06/01/2017	9045C
06/06/2017	Solid	I3444-08	Corrosivity	Cool 4 deg C	PSEG05	J13	BU-RO-298-060117	06/01/2017	9045C
06/06/2017	Solid	I3449-05	Corrosivity	Cool 4 deg C	PSEG01	K43	CARSONS-STOCKPILE(COM	06/02/2017	9045C

Date/Time 06/02/17 11:50 AM

Received by: JP

Relinquished by: Hassan

Date/Time 06/02/17 6:35 PM

Received by: AL

Relinquished by: JP