



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87825

Temp.: 20.4 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/02/2017	16:10
2	CAL2	1	Water			7.00	06/02/2017	16:11
3	CAL3	1	Water			10.02	06/02/2017	16:12
4	ICV	1	Water			7.01	06/02/2017	16:13
5	CCV1	1	Water			2.02	06/02/2017	16:15
6	I3450-02	1	Water			7.61	06/02/2017	16:17
7	I3450-02DUP	1	Water			7.62	06/02/2017	16:18
8	CCV2	1	Water			12.01	06/02/2017	16:20

2387825

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	CORROSIVITY I3450	WorkList ID :	99139	Date :	6/2/2017 3:51:05 PM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3450-02	Corrosivity	Cool 4 deg C	PSEG01	K43	A-2186	06/02/2017	9040C

Date/Time 06/02/17 3:51 PM
Received by: JP
Relinquished by: Harrison Lane

Date/Time 06/02/17 6:58 PM
Received by: H/L
Relinquished by: JO