



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87832

Temp.: 19.4 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	06/02/2017	17:10
2	CAL2	1	Water			7.01	06/02/2017	17:11
3	CAL3	1	Water			10.02	06/02/2017	17:13
4	ICV	1	Water			7.00	06/02/2017	17:15
5	CCV1	1	Water			2.02	06/02/2017	17:16
6	I3461-04	1	Water			7.21	06/02/2017	17:18
7	I3461-04DUP	1	Water			7.22	06/02/2017	17:20
8	CCV2	1	Water			12.01	06/02/2017	17:21

WORKLIST(Hardcopy Internal Chain)

1387832

WorkList Name : ph i3461

WorkList ID : 99142

Date : 6/2/2017 4:54:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3461-04✓	pH	Cool 4 deg C	CHEM05	K32	EFFLUENT-4	06/02/2017	SM 4500-PH B

Date/Time 06/02/17 5:00 PM
 Received by: SP
 Relinquished by: CD

Date/Time 06/02/17 3:00 PM
 Received by: CD
 Relinquished by: SP