



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87841

Temp.: 23.5 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/02/2017	18:30
2	CAL2	1	Water			6.99	06/02/2017	18:31
3	CAL3	1	Water			10.03	06/02/2017	18:32
4	ICV	1	Water			7.01	06/02/2017	18:33
5	CCV1	1	Water			2.01	06/02/2017	18:35
6	I3462-01	1	Solid	20.02	20	5.83	06/02/2017	18:36
7	I3462-01DUP	1	Solid	20.01	20	5.84	06/02/2017	18:37
8	CCV2	1	Water			12.01	06/02/2017	18:39

2387841

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	ph i3462	WorkList ID :	99145	Date :	6/2/2017 6:27:27 PM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3462-01	pH	Cool 4 deg C	TULL02	K22	FK-PS-01-005	06/02/2017	9045C

Date/Time 06/02/17 6:30 PM
Received by: JP
Relinquished by: SR

Date/Time 06/02/17 7:35 PM
Received by: CP
Relinquished by: JP