



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB87845
Analysis Date: 06/05/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 06/05/2017
Extraction IN Time: 08:00
Extraction OUT Time: 08:45

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB87845BL	LB87845BL	WATER	2	1000	100	3.0741	3.0741	3.02	3.0741	3.0741	0	0
2	LB87845BS	LB87845BS	WATER	2	1000	100	3.9618	3.9618	3.01	3.9807	3.9807	0.0189	18.9
3	LB87845BSD	LB87845BSD	WATER	2	1000	100	3.0522	3.0522	3.02	3.0712	3.0712	0.019	19
4	I3450-01	A-2186	WATER	2	1000	100	3.1847	3.1847	3.04	3.1852	3.1852	0.0005	0.5



QC Batch# LB87845 Test: TPH Analysis Date: 06/05/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2243
pH Paper 0-14	W2186
Sodium Sulfate	EP1735
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP52823
LCSWD	10.00 ML	WP54439
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 71 Dessicator Time In : 10:00
1.0000 gram Balance: 1.0003 In Time: 09:20
Bal Check Time: 08:25

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 70 Dessicator Time Out: 11:00
1.0000 gram Balance: 1.0005 Out Time: 10:20
Bal Check Time: 11:30

1887865

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	NON POPLAR i3450	WorkList ID :	99148	Date :	6/5/2017 7:48:13 AM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3450-01-11	TPH	Conc H2SO4 1c	PSEG01	K43	A-2186	06/02/2017	1664A

Date/Time 06/05/17 7:41 AM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 06/03/17 10:00 AM
Received by: J.S.
Relinquished by: 30