


245.1 LB87848
INSTRUMENT ID: CV1

Linear

μ Abs.:

33702



A= 0.0000e+000

B= 2.9574e-004

C= -6.6601e-002

Rho= 0.9996324

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.080	-0.080	-46	0.000	-46				
0.2	0.200	0.183	-0.017	843	0.0 %	843				
2.5	2.500	2.646	0.146	9172	0.0 %	9172				
5.0	5.000	4.924	-0.076	16874	0.0 %	16874				
7.5	7.500	7.628	0.128	26017	0.0 %	26017				
10.0	10.000	9.900	-0.100	33702	0.0 %	33702				

LB87848 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 05 Jun 2017 10:07:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-46	-	0.0000	245.1	PPB	05 Jun 2017 10:35:25	S	Std
0.2 - 1	50.2	843	-	0.2000	245.1	PPB	05 Jun 2017 10:37:33	S	Std
2.5 - 1	52.5	9172	-	2.5000	245.1	PPB	05 Jun 2017 10:39:41	S	Std
5.0 - 1	55	16874	-	5.0000	245.1	PPB	05 Jun 2017 10:41:49	S	Std
7.5 - 1	57.5	26017	-	7.5000	245.1	PPB	05 Jun 2017 10:43:57	S	Std
10.0 - 1	57.5	33702	-	10.0000	245.1	PPB	05 Jun 2017 10:46:04	S	Std
ICV82 - 1	ICV82	13828	4.0229	-245.1		PPB	05 Jun 2017 10:48:52	U	SMPL
ICB82 - 1	ICB82	22	-0.0601	-245.1		PPB	05 Jun 2017 10:55:58	U	SMPL
CCV59 - 1	CCV59	16908	4.9337	-245.1		PPB	05 Jun 2017 10:58:04	U	SMPL
CCB59 - 1	CCB59	-80	-0.0903	-245.1		PPB	05 Jun 2017 11:00:11	U	SMPL
CRA - 1	CRA	833	0.1797	-245.1		PPB	05 Jun 2017 11:02:17	U	SMPL
HighStd - 1	HighStd	35311	10.3762	-245.1		PPB	05 Jun 2017 11:04:24	U	SMPL
ChkStd - 1	ChkStd	25395	7.4436	-245.1		PPB	05 Jun 2017 11:06:31	U	SMPL
PB99435BL - 1	PBW	-13	-0.0704	-245.1		PPB	05 Jun 2017 11:11:19	U	SMPL
PB99435BS - 1	LCSW	14044	4.0867	-245.1		PPB	05 Jun 2017 11:16:24	U	SMPL
I3414-01 - 1	TE-PMW-11S	1732	0.4456	-245.1		PPB	05 Jun 2017 11:18:31	U	SMPL
I3414-01DUP - 1	TE-PMW-11SDUP	1523	0.3838	-245.1		PPB	05 Jun 2017 11:20:39	U	SMPL
I3414-01MS - 1	TE-PMW-11SMS	16365	4.7731	-245.1		PPB	05 Jun 2017 11:22:47	U	SMPL
I3414-01MSD - 1	TE-PMW-11SMSD	16862	4.9201	-245.1		PPB	05 Jun 2017 11:24:56	U	SMPL
CCV60 - 1	CCV60	16728	4.8805	-245.1		PPB	05 Jun 2017 11:27:03	U	SMPL
CCB60 - 1	CCB60	-128	-0.1045	-245.1		PPB	05 Jun 2017 11:29:09	U	SMPL
I3414-01LX5 - 1		185	-0.0119	-245.1		PPB	05 Jun 2017 11:31:16	U	SMPL
I3414-01A - 1		16816	4.9065	-245.1		PPB	05 Jun 2017 11:33:23	U	SMPL
CCV61 - 1	CCV61	16778	4.8953	-245.1		PPB	05 Jun 2017 11:35:29	U	SMPL
CCB61 - 1	CCB61	-57	-0.0835	-245.1		PPB	05 Jun 2017 11:37:36	U	SMPL