


7471A

LB87864
INSTRUMENT ID: CY1

Linear

 μ Abs.:

45150



A= 0.0000e+000

B= 2.2236e-004

C= -1.3517e-002

Rho= 0.9999530

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	96	0.000	96				
0.2	0.200	0.210	0.010	1005	0.0 %	1005				
2.5	2.500	2.464	-0.036	11142	0.0 %	11142				
5.0	5.000	5.049	0.049	22765	0.0 %	22765				
7.5	7.500	7.444	-0.056	33536	0.0 %	33536				
10.0	10.000	10.026	0.026	45150	0.0 %	45150				

LB87864 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 05 Jun 2017 15:46:19				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	96	-	0.00007471A	PPB		05 Jun 2017 15:52:47	S Std
0.2 - 1	50.2	1005	-	0.20007471A	PPB		05 Jun 2017 15:55:04	S Std
2.5 - 1	52.5	11142	-	2.50007471A	PPB		05 Jun 2017 15:57:22	S Std
5.0 - 1	55	22765	-	5.00007471A	PPB		05 Jun 2017 15:59:39	S Std
7.5 - 1	57.5	33536	-	7.50007471A	PPB		05 Jun 2017 16:01:57	S Std
10.0 - 1	510	45150	-	10.00007471A	PPB		05 Jun 2017 16:04:13	S Std
ICV84 - 1	ICV84	17376	3.8502	-7471A	PPB		05 Jun 2017 16:07:18	U SMPL
ICB84 - 1	ICB84	56	-0.0011	-7471A	PPB		05 Jun 2017 16:09:33	U SMPL
CCV65 - 1	CCV65	22458	4.9803	-7471A	PPB		05 Jun 2017 16:14:20	U SMPL
CCB65 - 1	CCB65	18	-0.0095	-7471A	PPB		05 Jun 2017 16:16:36	U SMPL
CRA - 1	CRA	1140	0.2400	-7471A	PPB		05 Jun 2017 16:18:51	U SMPL
HighStd - 1	HighStd	45207	10.0388	-7471A	PPB		05 Jun 2017 16:21:07	U SMPL
ChkStd - 1	ChkStd	30644	6.8005	-7471A	PPB		05 Jun 2017 16:23:23	U SMPL
PB99452BL - 1	PBS	197	0.0303	-7471A	PPB		05 Jun 2017 16:25:40	U SMPL
PB99452BS - 1	LCSS	14765	3.2696	-7471A	PPB		05 Jun 2017 16:27:57	U SMPL
I3437-01 - 1	COMP-1	508	0.0994	-7471A	PPB		05 Jun 2017 16:30:14	U SMPL
I3438-01 - 1	TP-5-6	970	0.2022	-7471A	PPB		05 Jun 2017 16:32:31	U SMPL
I3438-06 - 1	TP-7	1009	0.2108	-7471A	PPB		05 Jun 2017 16:34:48	U SMPL
I3438-11 - 1	TP-8-9	965	0.2011	-7471A	PPB		05 Jun 2017 16:37:07	U SMPL
I3440-01 - 1	EO-02-RO-4146-060117	3025	0.6591	-7471A	PPB		05 Jun 2017 16:39:23	U SMPL
CCV66 - 1	CCV66	21230	4.7072	-7471A	PPB		05 Jun 2017 16:41:39	U SMPL
CCB66 - 1	CCB66	91	0.0067	-7471A	PPB		05 Jun 2017 16:43:55	U SMPL
I3440-03 - 1	EO-02-RO-5078-060117	1683	0.3607	-7471A	PPB		05 Jun 2017 16:46:11	U SMPL
I3440-05 - 1	EO-03-RO-6703-060117	6696	1.4754	-7471A	PPB		05 Jun 2017 16:48:28	U SMPL
I3441-01 - 1	OR-RO-5267-060117	3108	0.6776	-7471A	PPB		05 Jun 2017 16:50:43	U SMPL
I3441-03 - 1	OR-RO-5004-060117	1735	0.3723	-7471A	PPB		05 Jun 2017 16:53:00	U SMPL
I3441-05 - 1	OR-RO-141-060117	1769	0.3798	-7471A	PPB		05 Jun 2017 16:55:16	U SMPL
I3441-07 - 1	OR-RO-6710-060117	2083	0.4497	-7471A	PPB		05 Jun 2017 16:57:33	U SMPL
I3441-09 - 1	OR-RO-6789-060117	3749	0.8201	-7471A	PPB		05 Jun 2017 16:59:50	U SMPL
I3441-11 - 1	OR-RO-17273-060117	1707	0.3661	-7471A	PPB		05 Jun 2017 17:02:08	U SMPL
I3442-01 - 1	OK-02-060117	91459	20.3234	-7471A	PPB		05 Jun 2017 17:04:25	U SMPL
I3443-01 - 1	SA-RO-6699AA-060117	330	0.0599	-7471A	PPB		05 Jun 2017 17:06:41	U SMPL
CCV67 - 1	CCV67	21762	4.8255	-7471A	PPB		05 Jun 2017 17:08:58	U SMPL
CCB67 - 1	CCB67	26	-0.0077	-7471A	PPB		05 Jun 2017 17:11:14	U SMPL
I3444-01 - 1	BU-RO-238-060117	787	0.1615	-7471A	PPB		05 Jun 2017 17:13:30	U SMPL
I3444-03 - 1	BU-RO-240-060117	1289	0.2731	-7471A	PPB		05 Jun 2017 17:15:46	U SMPL
I3444-05 - 1	BU-RO-286-060117	1273	0.2695	-7471A	PPB		05 Jun 2017 17:18:03	U SMPL
I3444-07 - 1	BU-RO-298-060117	819	0.1686	-7471A	PPB		05 Jun 2017 17:20:19	U SMPL
I3444-07DUP - 1	BU-RO-298-060117DUP	887	0.1837	-7471A	PPB		05 Jun 2017 17:22:35	U SMPL
I3444-07MS - 1	BU-RO-298-060117MS	15832	3.5069	-7471A	PPB		05 Jun 2017 17:24:52	U SMPL
I3444-07MSD - 1	BU-RO-298-060117MSD	15835	3.5076	-7471A	PPB		05 Jun 2017 17:27:09	U SMPL
I3449-01 - 1	CARSONS-STOCKPILE(COMP)	33187	7.3660	-7471A	PPB		05 Jun 2017 17:29:27	U SMPL
I3444-07LX5 - 1		175	0.0254	-7471A	PPB		05 Jun 2017 17:31:44	U SMPL
I3444-07A - 1		18274	4.0499	-7471A	PPB		05 Jun 2017 17:34:00	U SMPL
CCV68 - 1	CCV68	21991	4.8764	-7471A	PPB		05 Jun 2017 17:36:17	U SMPL
CCB68 - 1	CCB68	14	-0.0104	-7471A	PPB		05 Jun 2017 17:38:33	U SMPL
I3442-01DLX5 - 1		18564	4.1144	-7471A	PPB		05 Jun 2017 17:40:49	U SMPL
CCV69 - 1	CCV69	22736	5.0421	-7471A	PPB		05 Jun 2017 17:43:06	U SMPL
CCB69 - 1	CCB69	16	-0.0100	-7471A	PPB		05 Jun 2017 17:45:22	U SMPL