



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87901

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/06/2017	15:00
2	CAL2	1	Water			7.00	06/06/2017	15:01
3	CAL3	1	Water			10.02	06/06/2017	15:03
4	ICV	1	Water			7.01	06/06/2017	15:05
5	CCV1	1	Water			2.02	06/06/2017	15:06
6	I3486-03	1	Solid	20.02	20	8.28	06/06/2017	15:09
7	I3487-03	1	Solid	20.03	20	8.47	06/06/2017	15:10
8	I3488-03	1	Solid	20.04	20	7.13	06/06/2017	15:11
9	I3488-06	1	Solid	20.02	20	8.55	06/06/2017	15:12
10	I3488-09	1	Solid	20.05	20	8.04	06/06/2017	15:15
11	I3488-12	1	Solid	20.01	20	8.73	06/06/2017	15:16
12	I3488-15	1	Solid	20.02	20	9.03	06/06/2017	15:17
13	I3500-03	1	Solid	20.00	20	8.60	06/06/2017	15:19
14	I3506-02	1	Solid	20.03	20	7.02	06/06/2017	15:22
15	I3506-02DUP	1	Solid	20.04	20	7.03	06/06/2017	15:25
16	CCV2	1	Water			12.01	06/06/2017	15:26

WORKLIST(Hardcopy Internal Chain)

LB87901

WorkList Name : CORROSIONITY I3486

WorkList ID : 99248

Date : 6/6/2017 12:41:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/08/2017	Solid	I3486-03	Corrosivity	Cool 4 deg C	PSEG05	K51	HD-RO-20397-060517	06/05/2017	9045C
06/08/2017	Solid	I3487-03	Corrosivity	Cool 4 deg C	PSEG05	K61	OK-03-060517	06/05/2017	9045C
06/08/2017	Solid	I3488-03	Corrosivity	Cool 4 deg C	PSEG05	K51	TR-04-060517-A	06/05/2017	9045C
06/08/2017	Solid	I3488-06	Corrosivity	Cool 4 deg C	PSEG05	K51	TR-04-060517-B	06/05/2017	9045C
06/08/2017	Solid	I3488-09	Corrosivity	Cool 4 deg C	PSEG05	K51	TR-04-060517-C	06/05/2017	9045C
06/08/2017	Solid	I3488-12	Corrosivity	Cool 4 deg C	PSEG05	K51	TR-RO-211757-060517	06/05/2017	9045C
06/08/2017	Solid	I3488-15	Corrosivity	Cool 4 deg C	PSEG05	K51	TR-RO-45-060517	06/05/2017	9045C
06/09/2017	Solid	I3500-03	Corrosivity	Cool 4 deg C	PSEG05	K61	OR-01-060617	06/06/2017	9045C
06/13/2017	Solid	I3506-02	Corrosivity	Cool 4 deg C	PSEG01	K61	TP-1	06/06/2017	9045C

Date/Time 06/06/17 12:50 PM
 Received by: JS
 Relinquished by: JS

Date/Time 06/06/17 6:00 PM
 Received by: JS
 Relinquished by: JP