



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87902

Temp.: 23.7 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/06/2017	16:40
2	CAL2	1	Water			7.01	06/06/2017	16:42
3	CAL3	1	Water			10.03	06/06/2017	16:43
4	ICV	1	Water			6.99	06/06/2017	16:45
5	CCV1	1	Water			2.01	06/06/2017	16:46
6	I3503-01	1	Solid	20.02	20	8.36	06/06/2017	16:47
7	I3503-01DUP	1	Solid	20.01	20	8.37	06/06/2017	16:48
8	I3503-03	1	Solid	20.04	20	8.33	06/06/2017	16:49
9	I3503-05	1	Solid	20.05	20	7.87	06/06/2017	16:50
10	CCV2	1	Water			12.02	06/06/2017	16:51

LB87902

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph I3503

WorkList ID : 99253

Date : 6/6/2017 4:31:15 pm

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/06/2017	Solid	I3503-01	pH	Cool 4 deg C	UNIO02		PIER-53-A-COMP	06/06/2017	9045C
06/06/2017	Solid	I3503-03	pH	Cool 4 deg C	UNIO02		PIER-53-B-COMP	06/06/2017	9045C
06/06/2017	Solid	I3503-05	pH	Cool 4 deg C	UNIO02		PIER-56-COMP	06/06/2017	9045C

Date/Time 06/06/17 1:25 PM
Received by: SS
Relinquished by: SS

Date/Time 06/06/17 6:00 PM
Received by: SS
Relinquished by: SS