



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB87903

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	06/06/2017	13:15
2	I3414-01		1	>100.0	>212.0	06/06/2017	13:45
3	I3414-01DUP		1	>100.0	>212.0	06/06/2017	14:10
4	I3450-02		1	>100.0	>212.0	06/06/2017	14:40
5	I3469-09		1	>100.0	>212.0	06/06/2017	15:05

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 87903

Worklist Name : FP6617

Worklist ID : 99255

Date : 6/6/2017 12:30PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3414-01	Flash Point	Cool 4 deg C	TECT01	J51	TE-PMW-11S	05/31/2017	1010A
	Water	I3450-02	Flash Point	Cool 4 deg C	PSEG01	K43	A-2186	06/02/2017	1010A
	Water	I3469-09	Flash Point	Cool 4 deg C	AKRF02	K51	TW-1-20170603	06/03/2017	1010A

Date/Time

6/6/17 12:40 PM

Received by:

KS

Relinquished by:

SR

Date/Time

6/6/17 15:10

Received by:

SR

Relinquished by:

KS