



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87912

Temp.: 20.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/07/2017	12:06
2	CAL2	1	Water			7.00	06/07/2017	12:09
3	CAL3	1	Water			10.03	06/07/2017	12:11
4	ICV	1	Water			7.01	06/07/2017	12:13
5	CCV1	1	Water			2.02	06/07/2017	12:15
6	I3512-01	1	Water			4.30	06/07/2017	12:16
7	I3516-01	1	Water			5.28	06/07/2017	12:17
8	I3516-01DUP	1	Water			5.29	06/07/2017	12:19
9	CCV2	1	Water			12.01	06/07/2017	12:20

20387912

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3516

WorkList ID : 99285

Date : 6/7/2017 11:46:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3512-01-17	pH	NONE	STER08	K62	POOLWATER	05/30/2017	SM 4500-PH B
	Water	I3516-01-17	pH	Cool 4 deg C	STER01	K63	POOLWATER	06/06/2017	SM 4500-PH B

Date/Time 06/07/17 11:46 AM
Received by: HL
Relinquished by: HL

Date/Time 06/07/17 11:00 AM
Received by: HL
Relinquished by: HL