



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 06/08/2017DATE & TIME IN: 06/07/2017 17:00OVEN TEMP: 98 °CRun Number: LB87913DATE & TIME OUT: 06/08/2017 12:50OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB87913BL	LB87913BL	93.4802	93.4802	100	93.4802	93.4802	93.4802	0	0
2	I3469-09	TW-1-20170603	97.4596	97.4596	100	97.5128	97.5128	97.5128	0.0532	532
3	I3469-09DUP	TW-1-20170603DUP	87.1338	87.1338	100	87.1875	87.1875	87.1875	0.0537	537

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

2158913

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS 13469

WorkList ID : 99277

Date : 6/7/2017 10:44:27 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3469-09-F	TS	Cool 4 deg C	AKRF02	K51	TW-1-20170603	06/03/2017	SM2540 B

Date/Time 06/07/17 3:20 PM
Received by: J0
Relinquished by: HL

Date/Time 06/07/17 5:40 PM
Received by: HL
Relinquished by: SA