

7471A


LB87922

INSTRUMENT ID: CY1

Linear

 μ Abs.:

42265



A= 0.0000e+000

B= 2.3622e-004

C= -2.4589e-002

Rho= 0.9995888

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	140	0.000	140				
0.2	0.200	0.222	0.022	1044	0.0 %	1044				
2.5	2.500	2.348	-0.152	10043	0.0 %	10043				
5.0	5.000	5.201	0.201	22121	0.0 %	22121				
7.5	7.500	7.462	-0.038	31693	0.0 %	31693				
10.0	10.000	9.959	-0.041	42265	0.0 %	42265				

LB87922 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 07 Jun 2017 15:54:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	140	-	0.00007471A	PPB		07 Jun 2017 16:01:04	S	Std
0.2 - 1	50.2	1044	-	0.20007471A	PPB		07 Jun 2017 16:03:21	S	Std
2.5 - 1	52.5	10043	-	2.50007471A	PPB		07 Jun 2017 16:05:39	S	Std
5.0 - 1	55	22121	-	5.00007471A	PPB		07 Jun 2017 16:07:56	S	Std
7.5 - 1	57.5	31693	-	7.50007471A	PPB		07 Jun 2017 16:10:13	S	Std
10.0 - 1	510	42265	-	10.00007471A	PPB		07 Jun 2017 16:15:03	S	Std
ICV88 - 1	ICV88	16791	3.9417	-7471A	PPB		07 Jun 2017 16:17:48	U	SMPL
ICB88 - 1	ICB88	133	0.0068	-7471A	PPB		07 Jun 2017 16:20:04	U	SMPL
CCV94 - 1	CCV94	21862	5.1396	-7471A	PPB		07 Jun 2017 16:22:20	U	SMPL
CCB94 - 1	CCB94	146	0.0099	-7471A	PPB		07 Jun 2017 16:24:35	U	SMPL
CRA - 1	CRA	1033	0.2194	-7471A	PPB		07 Jun 2017 16:26:51	U	SMPL
HighStd - 1	HighStd	42726	10.0680	-7471A	PPB		07 Jun 2017 16:29:07	U	SMPL
ChkStd - 1	ChkStd	30897	7.2738	-7471A	PPB		07 Jun 2017 16:31:23	U	SMPL
PB99542BL - 1	PBS	42	-0.0147	-7471A	PPB		07 Jun 2017 16:33:40	U	SMPL
PB99542BS - 1	LCSS	15339	3.5987	-7471A	PPB		07 Jun 2017 16:35:57	U	SMPL
I3481-01 - 1	SU-04-060517	1812	0.4034	-7471A	PPB		07 Jun 2017 16:38:13	U	SMPL
I3482-01 - 1	CL-01-060517-A	1781	0.3961	-7471A	PPB		07 Jun 2017 16:40:31	U	SMPL
I3482-04 - 1	CL-01-060517-B	1496	0.3288	-7471A	PPB		07 Jun 2017 16:42:48	U	SMPL
I3482-07 - 1	CL-01-060517-C	3482	0.7979	-7471A	PPB		07 Jun 2017 16:45:06	U	SMPL
I3482-10 - 1	CL-01-060517-D	2329	0.5256	-7471A	PPB		07 Jun 2017 16:47:22	U	SMPL
CCV95 - 1	CCV95	21963	5.1635	-7471A	PPB		07 Jun 2017 16:49:39	U	SMPL
CCB95 - 1	CCB95	142	0.0090	-7471A	PPB		07 Jun 2017 16:51:54	U	SMPL
I3482-13 - 1	CL-01-060517-E	1864	0.4157	-7471A	PPB		07 Jun 2017 16:54:10	U	SMPL
I3482-13DUP - 1	CL-01-060517-EDUP	1969	0.4405	-7471A	PPB		07 Jun 2017 16:56:27	U	SMPL
I3482-13MS - 1	CL-01-060517-EMS	17807	4.1817	-7471A	PPB		07 Jun 2017 16:58:43	U	SMPL
I3482-13MSD - 1	CL-01-060517-EMSD	17311	4.0646	-7471A	PPB		07 Jun 2017 17:00:59	U	SMPL
I3486-01 - 1	HD-RO-20397-060517	2533	0.5737	-7471A	PPB		07 Jun 2017 17:03:16	U	SMPL
I3487-01 - 1	OK-03-060517	879	0.1830	-7471A	PPB		07 Jun 2017 17:05:33	U	SMPL
I3488-01 - 1	TR-04-060517-A	1514	0.3330	-7471A	PPB		07 Jun 2017 17:07:50	U	SMPL
I3488-04 - 1	TR-04-060517-B	1210	0.2612	-7471A	PPB		07 Jun 2017 17:10:08	U	SMPL
I3488-07 - 1	TR-04-060517-C	1163	0.2501	-7471A	PPB		07 Jun 2017 17:12:26	U	SMPL
I3488-10 - 1	TR-RO-211757-060517	1155	0.2482	-7471A	PPB		07 Jun 2017 17:14:42	U	SMPL
CCV96 - 1	CCV96	22105	5.1970	-7471A	PPB		07 Jun 2017 17:16:58	U	SMPL
CCB96 - 1	CCB96	149	0.0106	-7471A	PPB		07 Jun 2017 17:19:14	U	SMPL
I3488-13 - 1	TR-RO-45-060517	796	0.1634	-7471A	PPB		07 Jun 2017 17:21:31	U	SMPL
I3495-01 - 1	US-01	15490	3.6344	-7471A	PPB		07 Jun 2017 17:23:47	U	SMPL
I3495-02 - 1	DS-01	5461	1.2654	-7471A	PPB		07 Jun 2017 17:26:03	U	SMPL
I3496-01 - 1	E2-S6-BASE	13848	3.2465	-7471A	PPB		07 Jun 2017 17:28:20	U	SMPL
I3496-02 - 1	E2-S6-WSW	22072	5.1892	-7471A	PPB		07 Jun 2017 17:30:37	U	SMPL
I3500-01 - 1	OR-01-060617	1268	0.2749	-7471A	PPB		07 Jun 2017 17:32:54	U	SMPL
I3506-01 - 1	TP-1	1225	0.2648	-7471A	PPB		07 Jun 2017 17:35:11	U	SMPL
I3482-13LX5 - 1		427	0.0763	-7471A	PPB		07 Jun 2017 17:37:29	U	SMPL
I3482-13A - 1		18567	4.3613	-7471A	PPB		07 Jun 2017 17:39:47	U	SMPL
CCV97 - 1	CCV97	22106	5.1972	-7471A	PPB		07 Jun 2017 17:42:03	U	SMPL
CCB97 - 1	CCB97	138	0.0080	-7471A	PPB		07 Jun 2017 17:44:19	U	SMPL