



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87932

Temp.: 23.7 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/07/2017	17:25
2	CAL2	1	Water			7.01	06/07/2017	17:26
3	CAL3	1	Water			10.04	06/07/2017	17:29
4	ICV	1	Water			7.01	06/07/2017	17:31
5	CCV1	1	Water			2.01	06/07/2017	17:33
6	I3526-03	1	Solid	20.02	20	6.75	06/07/2017	17:35
7	I3526-03DUP	1	Solid	20.01	20	6.76	06/07/2017	17:36
8	I3541-03	1	Solid	20.03	20	5.85	06/07/2017	17:37
9	I3541-06	1	Solid	20.02	20	10.18	06/07/2017	17:39
10	CCV2	1	Water			12.02	06/07/2017	17:40

2087932

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3526 WorkList ID : 99298 Date : 6/7/2017 5:17:37 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3526-03	Corrosivity	Cool 4 deg C	PSEG05	K61	BU-01-060717	06/07/2017	9045C
	Solid	I3541-03	Corrosivity	Cool 4 deg C	PSEG01	L52	RT-4215	06/07/2017	9045C
	Solid	I3541-06	Corrosivity	Cool 4 deg C	PSEG01	L52	RT-3460	06/07/2017	9045C

Date/Time 06/07/15 5:14PM
Received by: JS
Relinquished by: JS

Date/Time 06/07/15 7:00PM
Received by: JS
Relinquished by: JS