

7471A


LB87956

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

42163



A= 0.0000e+000

B= 2.3427e-004

C= 1.0460e-002

Rho= 0.9996660

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	38	0.000	38				
0.2	0.200	0.220	0.020	895	0.0 %	895				
2.5	2.500	2.376	-0.124	10097	0.0 %	10097				
5.0	5.000	5.041	0.041	21474	0.0 %	21474				
7.5	7.500	7.655	0.155	32633	0.0 %	32633				
10.0	10.000	9.888	-0.112	42163	0.0 %	42163				

LB87956 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 08 Jun 2017 15:58:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	38	-	0.00007471A	PPB		08 Jun 2017 16:02:50	S	Std
0.2 - 1	50.2	895	-	0.20007471A	PPB		08 Jun 2017 16:05:07	S	Std
2.5 - 1	52.5	10097	-	2.50007471A	PPB		08 Jun 2017 16:07:25	S	Std
5.0 - 1	55	21474	-	5.00007471A	PPB		08 Jun 2017 16:09:43	S	Std
7.5 - 1	57.5	32633	-	7.50007471A	PPB		08 Jun 2017 16:12:00	S	Std
10.0 - 1	510	42163	-	10.00007471A	PPB		08 Jun 2017 16:16:51	S	Std
ICV90 - 1	ICV90	16583	3.8954	-7471A	PPB		08 Jun 2017 16:19:33	U	SMPL
ICB90 - 1	ICB90	-21	0.0055	-7471A	PPB		08 Jun 2017 16:21:48	U	SMPL
CCV07 - 1	CCV07	21450	5.0356	-7471A	PPB		08 Jun 2017 16:24:04	U	SMPL
CCB07 - 1	CCB07	-21	0.0055	-7471A	PPB		08 Jun 2017 16:26:20	U	SMPL
CRA - 1	CRA	889	0.2187	-7471A	PPB		08 Jun 2017 16:28:36	U	SMPL
HighStd - 1	HighStd	42249	9.9082	-7471A	PPB		08 Jun 2017 16:30:51	U	SMPL
ChkStd - 1	ChkStd	29999	7.0383	-7471A	PPB		08 Jun 2017 16:33:08	U	SMPL
PB99593BL - 1	PBS	-83	-0.0090	-7471A	PPB		08 Jun 2017 16:35:24	U	SMPL
PB99593BS - 1	LCSS	14746	3.4650	-7471A	PPB		08 Jun 2017 16:37:41	U	SMPL
I3503-01 - 1	PIER-53-A-COMP	28733	6.7418	-7471A	PPB		08 Jun 2017 16:39:58	U	SMPL
I3503-03 - 1	PIER-53-B-COMP	31344	7.3534	-7471A	PPB		08 Jun 2017 16:42:15	U	SMPL
I3503-05 - 1	PIER-56-COMP	35516	8.3308	-7471A	PPB		08 Jun 2017 16:44:33	U	SMPL
I3526-01 - 1	BU-01-060717	883	0.2173	-7471A	PPB		08 Jun 2017 16:46:51	U	SMPL
CCV08 - 1	CCV08	21510	5.0496	-7471A	PPB		08 Jun 2017 16:51:23	U	SMPL
CCB08 - 1	CCB08	-11	0.0079	-7471A	PPB		08 Jun 2017 16:53:39	U	SMPL
I3535-01 - 1	SOILSAMPLE-1	2176	0.5202	-7471A	PPB		08 Jun 2017 16:59:57	U	SMPL
I3535-01DUP - 1	SOILSAMPLE-1DUP	2126	0.5085	-7471A	PPB		08 Jun 2017 17:02:13	U	SMPL
I3535-01MS - 1	SOILSAMPLE-1MS	16345	3.8396	-7471A	PPB		08 Jun 2017 17:04:29	U	SMPL
I3535-01MSD - 1	SOILSAMPLE-1MSD	16457	3.8659	-7471A	PPB		08 Jun 2017 17:06:45	U	SMPL
I3541-01 - 1	RT-4215	1157	0.2815	-7471A	PPB		08 Jun 2017 17:09:02	U	SMPL
I3541-04 - 1	RT-3460	1046	0.2555	-7471A	PPB		08 Jun 2017 17:11:19	U	SMPL
I3559-15 - 1	BACKFILL-01	35	0.0187	-7471A	PPB		08 Jun 2017 17:13:36	U	SMPL
I3563-01 - 1	OK-04-060817	664	0.1660	-7471A	PPB		08 Jun 2017 17:15:53	U	SMPL
I3535-01LX5 - 1		424	0.1098	-7471A	PPB		08 Jun 2017 17:18:11	U	SMPL
I3535-01A - 1		19207	4.5101	-7471A	PPB		08 Jun 2017 17:20:28	U	SMPL
CCV09 - 1	CCV09	21732	5.1016	-7471A	PPB		08 Jun 2017 17:22:44	U	SMPL
CCB09 - 1	CCB09	-6	0.0091	-7471A	PPB		08 Jun 2017 17:25:00	U	SMPL