



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87957

Temp.: 21.3 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/08/2017	16:15
2	CAL2	1	Water			7.00	06/08/2017	16:16
3	CAL3	1	Water			10.03	06/08/2017	16:19
4	ICV	1	Water			7.00	06/08/2017	16:20
5	CCV1	1	Water			2.01	06/08/2017	16:21
6	I3561-01	1	Water			11.80	06/08/2017	16:22
7	I3561-01DUP	1	Water			11.81	06/08/2017	16:23
8	I3561-03	1	Water			12.15	06/08/2017	16:25
9	CCV2	1	Water			12.01	06/08/2017	16:27

20587437

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I3560 WorkList ID : 99357 Date : 6/8/2017 3:58:18 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3561-01	Corrosivity	Cool 4 deg C	SEAC01	L22	PURGE-WATER-NRL-65	06/07/2017	9040C
	Water	I3561-03	Corrosivity	Cool 4 deg C	SEAC01	L22	PURGE-WATER-NRL-74	06/07/2017	9040C

Date/Time 06/08/17 6:00 PM
Received by: JP
Relinquished by: JS

Date/Time 06/08/17 6:00 PM
Received by: CR
Relinquished by: JP