



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87962

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/08/2017	16:35
2	CAL2	1	Water			7.00	06/09/2017	16:36
3	CAL3	1	Water			10.02	06/08/2017	16:37
4	ICV	1	Water			6.99	06/08/2017	16:39
5	CCV1	1	Water			2.01	06/08/2017	16:40
6	I3567-01	1	Solid	20.02	20	6.35	06/08/2017	16:41
7	I3567-01DUP	1	Solid	20.03	20	6.36	06/08/2017	16:43
8	CCV2	1	Water			12.01	06/08/2017	16:45

2387962

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3567

WorkList ID : 99358

Date : 6/8/2017 4:15:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3567-01	pH	Cool 4 deg C	TULL02	M52	FK-PS-01-006	06/08/2017	9045C

Date/Time 06/08/17 11:00 AM
Received by: JP
Relinquished by: JP

Date/Time 06/08/17 6:00 PM
Received by: CJ
Relinquished by: JP