


7470A

LB87971
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35447



A= 0.0000e+000

B= 2.7815e-004

C= 1.7779e-002

Rho= 0.9995556

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	2	0.000	2				
0.2	0.200	0.224	0.024	741	0.0 %	741				
2.5	2.500	2.337	-0.163	8337	0.0 %	8337				
5.0	5.000	5.108	0.108	18301	0.0 %	18301				
7.5	7.500	7.635	0.135	27386	0.0 %	27386				
10.0	10.000	9.877	-0.123	35447	0.0 %	35447				

LB87971 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Jun 2017 09:28:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	2	-	0.0000	7470A	PPB	09 Jun 2017 09:28:47	S	Std
0.2 - 1	50.2	741	-	0.2000	7470A	PPB	09 Jun 2017 09:30:56	S	Std
2.5 - 1	52.5	8337	-	2.5000	7470A	PPB	09 Jun 2017 09:33:04	S	Std
5.0 - 1	55	18301	-	5.0000	7470A	PPB	09 Jun 2017 09:35:12	S	Std
7.5 - 1	57.5	27386	-	7.5000	7470A	PPB	09 Jun 2017 09:37:20	S	Std
10.0 - 1	510	35447	-	10.0000	7470A	PPB	09 Jun 2017 09:39:28	S	Std
ICV91 - 1	ICV91	14518	4.0560	-	7470A	PPB	09 Jun 2017 09:42:07	U	SMPL
ICB91 - 1	ICB91	-65	-0.0003	-	7470A	PPB	09 Jun 2017 09:44:14	U	SMPL
CCV10 - 1	CCV10	18436	5.1458	-	7470A	PPB	09 Jun 2017 09:46:20	U	SMPL
CCB10 - 1	CCB10	-69	-0.0014	-	7470A	PPB	09 Jun 2017 09:48:27	U	SMPL
CRA - 1	CRA	735	0.2222	-	7470A	PPB	09 Jun 2017 09:50:33	U	SMPL
HighStd - 1	HighStd	35527	9.8997	-	7470A	PPB	09 Jun 2017 09:52:40	U	SMPL
ChkStd - 1	ChkStd	25104	7.0005	-	7470A	PPB	09 Jun 2017 09:54:47	U	SMPL
PB99601BL - 1	PBW	-84	-0.0056	-	7470A	PPB	09 Jun 2017 09:56:54	U	SMPL
PB99601BS - 1	LCSW	13414	3.7489	-	7470A	PPB	09 Jun 2017 09:59:02	U	SMPL
PB99601TB - 1	PB99568TB	-49	0.0041	-	7470A	PPB	09 Jun 2017 10:01:09	U	SMPL
I3526-03 - 1	BU-01-060717	-19	0.0125	-	7470A	PPB	09 Jun 2017 10:03:17	U	SMPL
I3541-03 - 1	RT-4215	6	0.0194	-	7470A	PPB	09 Jun 2017 10:05:26	U	SMPL
I3541-06 - 1	RT-3460	-7	0.0158	-	7470A	PPB	09 Jun 2017 10:07:34	U	SMPL
I3541-06DUP - 1	RT-3460DUP	-12	0.0144	-	7470A	PPB	09 Jun 2017 10:09:41	U	SMPL
CCV11 - 1	CCV11	18702	5.2198	-	7470A	PPB	09 Jun 2017 10:11:48	U	SMPL
CCB11 - 1	CCB11	-68	-0.0011	-	7470A	PPB	09 Jun 2017 10:13:55	U	SMPL
I3541-06MS - 1	RT-3460MS	13193	3.6875	-	7470A	PPB	09 Jun 2017 10:16:01	U	SMPL
I3541-06MSD - 1	RT-3460MSD	13399	3.7448	-	7470A	PPB	09 Jun 2017 10:18:08	U	SMPL
I3541-06LX5 - 1		-70	-0.0017	-	7470A	PPB	09 Jun 2017 10:20:15	U	SMPL
I3541-06A - 1		14386	4.0193	-	7470A	PPB	09 Jun 2017 10:22:22	U	SMPL
CCV12 - 1	CCV12	18433	5.1450	-	7470A	PPB	09 Jun 2017 10:24:29	U	SMPL
CCB12 - 1	CCB12	-73	-0.0025	-	7470A	PPB	09 Jun 2017 10:26:37	U	SMPL