

LB87975


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

32157



A= 0.0000e+000

B= 3.0947e-004

C= -3.1148e-003

Rho= 0.9998415

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-13	0.000	-13				
0.2	0.200	0.205	0.005	671	0.0 %	671				
2.5	2.500	2.419	-0.081	7827	0.0 %	7827				
5.0	5.000	5.129	0.129	16583	0.0 %	16583				
7.5	7.500	7.506	0.006	24266	0.0 %	24266				
10.0	10.000	9.948	-0.052	32157	0.0 %	32157				

LB87975 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Jun 2017 11:00:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-13	-	0.0000	245.1	PPB	09 Jun 2017 11:08:23	S	Std
0.2 - 1	50.2	671	-	0.2000	245.1	PPB	09 Jun 2017 11:10:31	S	Std
2.5 - 1	52.5	7827	-	2.5000	245.1	PPB	09 Jun 2017 11:12:40	S	Std
5.0 - 1	55	16583	-	5.0000	245.1	PPB	09 Jun 2017 11:14:48	S	Std
7.5 - 1	57.5	24266	-	7.5000	245.1	PPB	09 Jun 2017 11:16:56	S	Std
10.0 - 1	510	32157	-	10.0000	245.1	PPB	09 Jun 2017 11:19:03	S	Std
ICV92 - 1	ICV92	12360	3.8219	-245.1		PPB	09 Jun 2017 11:21:39	U	SMPL
ICB92 - 1	ICB92	-60	-0.0217	-245.1		PPB	09 Jun 2017 11:23:46	U	SMPL
CCV13 - 1	CCV13	16561	5.1220	-245.1		PPB	09 Jun 2017 11:25:52	U	SMPL
CCB13 - 1	CCB13	-56	-0.0204	-245.1		PPB	09 Jun 2017 11:27:58	U	SMPL
CRA - 1	CRA	675	0.2058	-245.1		PPB	09 Jun 2017 11:30:05	U	SMPL
HighStd - 1	HighStd	31839	9.8500	-245.1		PPB	09 Jun 2017 11:32:12	U	SMPL
ChkStd - 1	ChkStd	23315	7.2121	-245.1		PPB	09 Jun 2017 11:34:19	U	SMPL
PB99602BL - 1	PBW	-87	-0.0300	-245.1		PPB	09 Jun 2017 11:36:27	U	SMPL
PB99602BS - 1	LCSW	13905	4.3000	-245.1		PPB	09 Jun 2017 11:38:35	U	SMPL
I3469-09 - 1	TW-1-20170603	29	0.0059	-245.1		PPB	09 Jun 2017 11:45:16	U	SMPL
I3469-09DUP - 1	TW-1-20170603DUP	0	-0.0031	-245.1		PPB	09 Jun 2017 11:47:24	U	SMPL
I3469-09MS - 1	TW-1-20170603MS	14555	4.5012	-245.1		PPB	09 Jun 2017 11:49:32	U	SMPL
I3469-09MSD - 1	TW-1-20170603MSD	14158	4.3783	-245.1		PPB	09 Jun 2017 11:51:41	U	SMPL
I3529-01 - 1	32220008-3	-78	-0.0273	-245.1		PPB	09 Jun 2017 11:53:48	U	SMPL
CCV14 - 1	CCV14	16318	5.0468	-245.1		PPB	09 Jun 2017 11:55:55	U	SMPL
CCB14 - 1	CCB14	-35	-0.0139	-245.1		PPB	09 Jun 2017 11:58:02	U	SMPL
I3538-01 - 1	COMP	424	0.1281	-245.1		PPB	09 Jun 2017 12:00:09	U	SMPL
I3469-09LX5 - 1		2	-0.0025	-245.1		PPB	09 Jun 2017 12:02:16	U	SMPL
I3469-09A - 1		15022	4.6457	-245.1		PPB	09 Jun 2017 12:04:23	U	SMPL
CCV15 - 1	CCV15	16487	5.0991	-245.1		PPB	09 Jun 2017 12:06:30	U	SMPL
CCB15 - 1	CCB15	-45	-0.0170	-245.1		PPB	09 Jun 2017 12:08:37	U	SMPL