

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB87976

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP55256
calibration std. phosphate 1 ppm	WP55520
calibration std. phosphate 0.5 ppm	WP55519
calibration std. phosphate 0.3 ppm	WP55518
calibration std. 0 ppm	WP55515
phosphate ICV std.	WP55521
phosphate CCV std.	WP55522
calibration std. phosphate 0.1 ppm	WP55517
calibration std. phosphate 0.05 ppm	WP55516
Combined reagent	WP55525
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749

Intercept: -0.0013Slope: 0.6507Regression: 0.999803

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	06/09/2017	11:06
2	CAL2	0.05	1	50	50	0.033	06/09/2017	11:06
3	CAL3	0.10	1	50	50	0.064	06/09/2017	11:07
4	CAL4	0.30	1	50	50	0.185	06/09/2017	11:07
5	CAL5	0.50	1	50	50	0.330	06/09/2017	11:08
6	CAL6	1.00	1	50	50	0.649	06/09/2017	11:08



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB87976

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.316	0.488	06/09/2017	11:09
2	ICB		1	50	50	0.000	0.002	06/09/2017	11:09
3	CCV1	0.50	1	50	50	0.329	0.508	06/09/2017	11:10
4	CCB1		1	50	50	0.000	0.002	06/09/2017	11:10
5	PB99591BL		1	50	50	0.001	0.004	06/09/2017	11:11
6	PB99591BS	0.50	1	50	50	0.313	0.483	06/09/2017	11:11
7	I3480-01		1	50	50	0.022	0.036	06/09/2017	11:12
8	I3480-01DUP		1	50	50	0.021	0.034	06/09/2017	11:12
9	I3480-01MS	0.50	1	50	50	0.360	0.555	06/09/2017	11:13
10	I3480-01MSD	0.50	1	50	50	0.357	0.551	06/09/2017	11:13
11	CCV2	0.50	1	50	50	0.327	0.505	06/09/2017	11:14
12	CCB2		1	50	50	0.000	0.002	06/09/2017	11:14