



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87981

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/09/2017	14:00
2	CAL2	1	Water			7.01	06/09/2017	14:01
3	CAL3	1	Water			10.03	06/09/2017	14:03
4	ICV	1	Water			6.99	06/09/2017	14:05
5	CCV1	1	Water			2.01	06/09/2017	14:06
6	I3576-06	1	Solid	20.02	20	5.88	06/09/2017	14:09
7	I3576-06DUP	1	Solid	20.03	20	5.89	06/09/2017	14:10
8	I3583-03	1	Solid	20.05	20	8.86	06/09/2017	14:11
9	I3583-06	1	Solid	20.02	20	7.95	06/09/2017	14:13
10	I3584-03	1	Solid	20.04	20	7.43	06/09/2017	14:15
11	I3586-02	1	Solid	20.03	20	10.93	06/09/2017	14:16
12	I3587-02	1	Solid	20.03	20	11.79	06/09/2017	14:19
13	I3592-03	1	Solid	20.00	20	8.06	06/09/2017	14:20
14	I3594-03	1	Solid	20.03	20	8.63	06/09/2017	14:21
15	I3594-06	1	Solid	20.05	20	8.51	06/09/2017	14:23
16	CCV2	1	Water			12.01	06/09/2017	14:26
17	I3594-09	1	Solid	20.03	20	7.81	06/09/2017	14:29
18	I3598-03	1	Solid	20.04	20	7.61	06/09/2017	14:30
19	I3598-07	1	Solid	20.05	20	7.18	06/09/2017	14:32
20	CCV3	1	Water			12.01	06/09/2017	14:34

21387981

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3576

WorkList ID : 99395

Date : 6/9/2017 12:37:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2017	Solid	I3576-06	Corrosivity	Cool 4 deg C	PSEG01	J31	TP-OGR1-OGR2	06/08/2017	9045C
06/13/2017	Solid	I3583-03	Corrosivity	Cool 4 deg C	PSEG01	J31	RO-2669	06/08/2017	9045C
06/13/2017	Solid	I3583-06	Corrosivity	Cool 4 deg C	PSEG01	J31	RO-2342	06/08/2017	9045C
06/13/2017	Solid	I3584-03	Corrosivity	Cool 4 deg C	PSEG01	J61	DUCT-BANK	06/08/2017	9045C
	Solid	I3586-02	Corrosivity	Cool 4 deg C	SEAC01	M51	CORE-SAMPLE-NRL-65	06/09/2017	9045C
	Solid	I3587-02	Corrosivity	Cool 4 deg C	SEAC01	M42	CORE-SAMPLE-NRL-74	06/09/2017	9045C
06/14/2017	Solid	I3592-03	Corrosivity	Cool 4 deg C	PSEG05	J61	SU-03-060917	06/09/2017	9045C
06/14/2017	Solid	I3594-03	Corrosivity	Cool 4 deg C	PSEG05	J62	EO-01-060917-A	06/09/2017	9045C
06/14/2017	Solid	I3594-06	Corrosivity	Cool 4 deg C	PSEG05	J62	EO-01-060917-B	06/09/2017	9045C
06/14/2017	Solid	I3594-09	Corrosivity	Cool 4 deg C	PSEG05	J62	EO-01-060917-C	06/09/2017	9045C
06/14/2017	Solid	I3598-03	Corrosivity	Cool 4 deg C	PSEG01	J63	VNJ-204-COMP	06/09/2017	9045C
06/14/2017	Solid	I3598-07	Corrosivity	Cool 4 deg C	PSEG01	J63	VNJ-226-COMP	06/09/2017	9045C

Date/Time 06/09/17 12:37:49

Received by: JH

Relinquished by: [Signature]

Date/Time 06/09/17 3:00 PM

Received by: HL

Relinquished by: JP