

Chemtech Consulting Group

Analytical Review Report

Date Printed : 6/13/17

Analyst : AK

Data File : LB87990.csv

Approved By : 

Approved Date : 6/14/17

Worksheet # :

m9056-A-Inorganic Anions-1B.

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Sulfate												
STD1	STD1				W	5/15/17						
Sulfate		PASS	0.000		0.000						mg/L	
STD2	STD2				W	5/15/17						
Sulfate		PASS	1.047		1.05						mg/L	
STD3	STD3				W	5/15/17						
Sulfate		PASS	2.987		2.99						mg/L	
STD4	STD4				W	5/15/17						
Sulfate		PASS	6.128		6.13						mg/L	
STD5	STD5				W	5/15/17						
Sulfate		PASS	14.359		14.4						mg/L	
STD6	STD6				W	5/15/17						
Sulfate		PASS	30.063		30.1						mg/L	
STD7	STD7				W	5/15/17						
Sulfate		PASS	37.666		37.7						mg/L	
ICV1	ICV1				W	5/15/17						
Sulfate		PASS	14.875		14.90	99.0	90	110			mg/L	
ICB1	ICB1				W	5/15/17						
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L	
CCV1	CCV1				W	6/9/17						
Sulfate		PASS	13.968		14.00	93.0	90	110			mg/L	
CCB1	CCB1				W	6/9/17						
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L	
LB87990BLS	LB87990BLS				S	6/9/17						
Sulfate		PASS	0.000		0.000			+/-0.0960			mg/Kg	
LB87990BSS	LB87990BSS				S	6/9/17						
Sulfate		PASS	13.825		276.00	92.0	90.00	110.00			mg/Kg	
I3604-01	B-2			1	S	6/9/17						
Sulfate		PASS	3.909		95.7						mg/Kg	
I3604-01D	B-2D			1	S	6/9/17						
Sulfate		PASS	3.939		96.70				1.0	20	mg/Kg	
I3604-01S	B-2S			1	S	6/9/17						
Sulfate		PASS	17.278		423.0	89.0	80	120			mg/Kg	
I3604-01SD	B-2SD			1	S	6/9/17						
Sulfate		PASS	17.378		425.0	90.0	80	120	0.5	15	mg/Kg	
CCV2	CCV2				W	6/9/17						
Sulfate		PASS	14.019		14.00	93.0	90	110			mg/L	
CCB2	CCB2				W	6/9/17						
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L	