



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88011

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/10/2017	12:20
2	CAL2	1	Water			7.00	06/10/2017	12:21
3	CAL3	1	Water			10.04	06/10/2017	12:22
4	ICV	1	Water			7.01	06/10/2017	12:23
5	CCV1	1	Water			2.01	06/10/2017	12:25
6	I3615-02	1	Solid	20.02	20	8.27	06/10/2017	12:30
7	I3615-06	1	Solid	20.04	20	8.92	06/10/2017	12:33
8	I3624-03	1	Solid	20.04	20	9.11	06/10/2017	12:35
9	I3624-06	1	Solid	20.01	20	8.96	06/10/2017	12:37
10	I3624-06DUP	1	Solid	20.03	20	8.97	06/10/2017	12:39
11	CCV2	1	Water			12.01	06/10/2017	12:40

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3615

WorkList ID : 99411

Date : 6/10/2017 10:21:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3615-02	Corrosivity	Cool 4 deg C	DEWB01	M42	HA4-WCA	06/09/2017	9045C
	Solid	I3615-06	Corrosivity	Cool 4 deg C	DEWB01	M42	HA4-WCB	06/09/2017	9045C
06/14/2017	Solid	I3624-03	Corrosivity	Cool 4 deg C	PSEG05	M52	NB-RO6151-060917	06/09/2017	9045C
06/14/2017	Solid	I3624-06	Corrosivity	Cool 4 deg C	PSEG05	M52	NB-RO6822-060917	06/09/2017	9045C

Date/Time 06-10-17 10:20 AM

Received by: 7H

Relinquished by: CR

Date/Time 06-10-17 10:20 AM

Received by: CR

Relinquished by: SD