


7470A

L988012
INSTRUMENT 1D: CV1

Linear

 μ Abs.:

38246



A= 0.0000e+000

B= 2.6456e-004

C= 4.7231e-002

Rho= 0.9994061

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.028	0.028	-74	0.000	-74				
0.2	0.200	0.228	0.028	685	0.0 %	685				
2.5	2.500	2.611	0.111	9690	0.0 %	9690				
5.0	5.000	4.817	-0.183	18028	0.0 %	18028				
7.5	7.500	7.351	-0.149	27608	0.0 %	27608				
10.0	10.000	10.165	0.165	38246	0.0 %	38246				

LB88012

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jun 2017 11:13:20

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	-74	-	0.00007470A	PPB	12 Jun 2017 11:35:23	S	Std
0.2 - 1	50.2	685	-	0.20007470A	PPB	12 Jun 2017 11:37:32	S	Std
2.5 - 1	52.5	9690	-	2.50007470A	PPB	12 Jun 2017 11:39:40	S	Std
5.0 - 1	55	18028	-	5.00007470A	PPB	12 Jun 2017 11:41:48	S	Std
7.5 - 1	57.5	27608	-	7.50007470A	PPB	12 Jun 2017 11:43:56	S	Std
10.0 - 1	59	38246	-	10.00007470A	PPB	12 Jun 2017 11:48:28	S	Std
ICV93 - 1	ICV93	13975	3.7444	-7470A	PPB	12 Jun 2017 11:53:07	U	SMPL
ICB93 - 1	ICB93	-90	0.0234	-7470A	PPB	12 Jun 2017 11:55:13	U	SMPL
CCV18 - 1	CCV18	18409	4.9174	-7470A	PPB	12 Jun 2017 11:57:20	U	SMPL
CCB18 - 1	CCB18	-96	0.0218	-7470A	PPB	12 Jun 2017 11:59:26	U	SMPL
CRA - 1	CRA	710	0.2351	-7470A	PPB	12 Jun 2017 12:01:33	U	SMPL
HighStd - 1	HighStd	38408	10.2083	-7470A	PPB	12 Jun 2017 12:03:40	U	SMPL
ChkStd - 1	ChkStd	25283	6.7360	-7470A	PPB	12 Jun 2017 12:05:47	U	SMPL
PB99636BL - 1	PBW	-118	0.0160	-7470A	PPB	12 Jun 2017 12:07:55	U	SMPL
PB99636BS - 1	LCSW	14672	3.9288	-7470A	PPB	12 Jun 2017 12:10:02	U	SMPL
PB99636TB - 1	PB99609TB	-87	0.0242	-7470A	PPB	12 Jun 2017 12:12:10	U	SMPL
I3560-01 - 1	FILTER-MEDIA	-5	0.0459	-7470A	PPB	12 Jun 2017 12:14:18	U	SMPL
I3561-02 - 1	SOIL-CUTTINGS-NRL-65	-36	0.0377	-7470A	PPB	12 Jun 2017 12:16:26	U	SMPL
I3563-03 - 1	OK-04-060817	-38	0.0372	-7470A	PPB	12 Jun 2017 12:18:35	U	SMPL
I3575-03 - 1	HR-06-060817-A	-39	0.0369	-7470A	PPB	12 Jun 2017 12:20:41	U	SMPL
CCV19 - 1	CCV19	17966	4.8002	-7470A	PPB	12 Jun 2017 12:22:49	U	SMPL
CCB19 - 1	CCB19	-92	0.0229	-7470A	PPB	12 Jun 2017 12:24:55	U	SMPL
I3575-06 - 1	HR-06-060817-B	-56	0.0324	-7470A	PPB	12 Jun 2017 12:27:02	U	SMPL
I3575-09 - 1	HR-06-060817-C	-52	0.0335	-7470A	PPB	12 Jun 2017 12:29:09	U	SMPL
I3575-09DUP - 1	HR-06-060817-CDUP	-58	0.0319	-7470A	PPB	12 Jun 2017 12:31:16	U	SMPL
I3575-09MS - 1	HR-06-060817-CMS	13929	3.7322	-7470A	PPB	12 Jun 2017 12:33:23	U	SMPL
I3575-09MSD - 1	HR-06-060817-CMSD	13959	3.7402	-7470A	PPB	12 Jun 2017 12:35:31	U	SMPL
PB99637BL - 1	PBW	-82	0.0255	-7470A	PPB	12 Jun 2017 12:37:38	U	SMPL
PB99637BS - 1	LCSW	14377	3.8507	-7470A	PPB	12 Jun 2017 12:39:46	U	SMPL
I3561-01 - 1	PURGE-WATER-NRL-65	937137	HIGH	-7470A	PPB	12 Jun 2017 12:41:55	X	
I3561-01DUP - 1	PURGE-WATER-NRL-65DUP	895638	HIGH	-7470A	PPB	12 Jun 2017 12:43:45	X	
I3561-01MS - 1	PURGE-WATER-NRL-65MS	849114	HIGH	-7470A	PPB	12 Jun 2017 12:45:50	X	
CCV20 - 1	CCV20	16870	4.5103	-7470A	PPB	12 Jun 2017 12:47:58	U	SMPL
CCB20 - 1	CCB20	-388	-0.0554	-7470A	PPB	12 Jun 2017 12:50:24	U	SMPL
I3561-01MSD - 1	PURGE-WATER-NRL-65MSD	838425	HIGH	-7470A	PPB	12 Jun 2017 12:52:31	X	
I3561-03 - 1	PURGE-WATER-NRL-74	157396	41.6872	-7470A	PPB	12 Jun 2017 12:54:19	U	SMPL
I3575-09LX5 - 1		-715	-0.1419	-7470A	PPB	12 Jun 2017 12:56:42	U	SMPL
I3575-09A - 1		13982	3.7462	-7470A	PPB	12 Jun 2017 12:58:51	U	SMPL
I3561-01LX5 - 1		348431	HIGH	-7470A	PPB	12 Jun 2017 13:00:59	X	
I3561-01A - 1		930852	HIGH	-7470A	PPB	12 Jun 2017 13:02:50	X	
I3561-01DLX10 - 1		224406	59.4150	-7470A	PPB	12 Jun 2017 13:04:42	U	SMPL
I3561-01DUPDLX10 - 1		209348	55.4314	-7470A	PPB	12 Jun 2017 13:06:57	U	SMPL
I3561-01MSDLX10 - 1		181785	48.1394	-7470A	PPB	12 Jun 2017 13:09:08	U	SMPL
I3561-01MSDDLX10 - 1		176338	46.6984	-7470A	PPB	12 Jun 2017 13:11:18	U	SMPL
CCV21 - 1	CCV21	17892	4.7807	-7470A	PPB	12 Jun 2017 13:13:28	U	SMPL
CCB21 - 1	CCB21	-223	-0.0118	-7470A	PPB	12 Jun 2017 13:15:37	U	SMPL
I3561-01LDLX50 - 1		52507	13.9382	-7470A	PPB	12 Jun 2017 13:17:44	U	SMPL
I3561-01ADLX10 - 1		247406	65.4998	-7470A	PPB	12 Jun 2017 13:19:52	U	SMPL
I3561-01DL2X100 - 1		26397	7.0307	-7470A	PPB	12 Jun 2017 13:22:00	U	SMPL
I3561-01DUPDL2X100 - 1		24322	6.4817	-7470A	PPB	12 Jun 2017 13:24:09	U	SMPL
I3561-01MSDL2X100 - 1		21118	5.6341	-7470A	PPB	12 Jun 2017 13:26:17	U	SMPL
I3561-01MSDDL2X100 - 1		20853	5.5640	-7470A	PPB	12 Jun 2017 13:28:25	U	SMPL
I3561-01LDL2X500 - 1		5442	1.4869	-7470A	PPB	12 Jun 2017 13:30:33	U	SMPL
I3561-01ADL2X100 - 1		26899	7.1635	-7470A	PPB	12 Jun 2017 13:32:41	U	SMPL
I3561-03DLX10 - 1		16997	4.5439	-7470A	PPB	12 Jun 2017 13:34:50	U	SMPL
CCV22 - 1	CCV22	18468	4.9330	-7470A	PPB	12 Jun 2017 13:36:57	U	SMPL
CCB22 - 1	CCB22	-97	0.0216	-7470A	PPB	12 Jun 2017 13:39:04	U	SMPL