

LB88017


7471A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48081



A= 0.0000e+000

B= 2.0849e-004

C= 2.7627e-003

Rho= 0.9999827

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	1	0.000	1				
0.2	0.200	0.212	0.012	1005	0.0 %	1005				
2.5	2.500	2.512	0.012	12034	0.0 %	12034				
5.0	5.000	4.960	-0.040	23777	0.0 %	23777				
7.5	7.500	7.486	-0.014	35893	0.0 %	35893				
10.0	10.000	10.027	0.027	48081	0.0 %	48081				

LB88017 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 12 Jun 2017 14:12:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1	-	0.0000	7471A	PPB	12 Jun 2017 14:42:06	S	Std
0.2 - 1	50.2	1005	-	0.2000	7471A	PPB	12 Jun 2017 14:44:24	S	Std
2.5 - 1	52.5	12034	-	2.5000	7471A	PPB	12 Jun 2017 14:46:42	S	Std
5.0 - 1	55	23777	-	5.0000	7471A	PPB	12 Jun 2017 14:48:59	S	Std
7.5 - 1	57.5	35893	-	7.5000	7471A	PPB	12 Jun 2017 14:51:17	S	Std
10.0 - 1	50	48081	-	10.0000	7471A	PPB	12 Jun 2017 14:56:11	S	Std
ICV94 - 1	ICV94	18235	3.8045	-7471A	PPB	12 Jun 2017 14:58:51	U	SMPL	
ICB94 - 1	ICB94	-78	-0.0135	-7471A	PPB	12 Jun 2017 15:01:06	U	SMPL	
CCV23 - 1	CCV23	23868	4.9789	-7471A	PPB	12 Jun 2017 15:03:22	U	SMPL	
CCB23 - 1	CCB23	-72	-0.0122	-7471A	PPB	12 Jun 2017 15:05:38	U	SMPL	
CRA - 1	CRA	976	0.2062	-7471A	PPB	12 Jun 2017 15:07:54	U	SMPL	
HighStd - 1	HighStd	45523	9.4937	-7471A	PPB	12 Jun 2017 15:10:09	U	SMPL	
ChkStd - 1	ChkStd	33880	7.0663	-7471A	PPB	12 Jun 2017 15:14:10	U	SMPL	
PB99648BL - 1	PBS	-148	-0.0281	-7471A	PPB	12 Jun 2017 15:16:26	U	SMPL	
PB99648BS - 1	LCSS	17257	3.6006	-7471A	PPB	12 Jun 2017 15:18:42	U	SMPL	
I3566-02 - 1	SB-01-COMP(0-10)	884	0.1871	-7471A	PPB	12 Jun 2017 15:21:00	U	SMPL	
I3566-04 - 1	SB-02-COMP(0-10)	3578	0.7487	-7471A	PPB	12 Jun 2017 15:23:17	U	SMPL	
I3566-06 - 1	SB-03-COMP(0-4.0)	18619	3.8846	-7471A	PPB	12 Jun 2017 15:25:34	U	SMPL	
I3566-08 - 1	SB-04-COMP(0-10)	1174	0.2475	-7471A	PPB	12 Jun 2017 15:27:52	U	SMPL	
I3566-10 - 1	SB-05-COMP(0-8.0)	613	0.1306	-7471A	PPB	12 Jun 2017 15:30:08	U	SMPL	
CCV24 - 1	CCV24	23765	4.9575	-7471A	PPB	12 Jun 2017 15:32:24	U	SMPL	
CCB24 - 1	CCB24	-51	-0.0079	-7471A	PPB	12 Jun 2017 15:34:40	U	SMPL	
I3566-12 - 1	SB-07-COMP(0-4.0)	3122	0.6537	-7471A	PPB	12 Jun 2017 15:36:56	U	SMPL	
I3566-14 - 1	SB-08-COMP(0-10)	6060	1.2662	-7471A	PPB	12 Jun 2017 15:39:12	U	SMPL	
I3567-01 - 1	FK-PS-01-006	555	0.1185	-7471A	PPB	12 Jun 2017 15:41:28	U	SMPL	
I3575-01 - 1	HR-06-060817-A	756	0.1604	-7471A	PPB	12 Jun 2017 15:43:45	U	SMPL	
I3575-01DUP - 1	HR-06-060817-ADUP	851	0.1802	-7471A	PPB	12 Jun 2017 15:46:01	U	SMPL	
I3575-01MS - 1	HR-06-060817-AMS	16435	3.4293	-7471A	PPB	12 Jun 2017 15:48:18	U	SMPL	
I3575-01MSD - 1	HR-06-060817-AMSD	16659	3.4760	-7471A	PPB	12 Jun 2017 15:50:35	U	SMPL	
I3575-04 - 1	HR-06-060817-B	4891	1.0225	-7471A	PPB	12 Jun 2017 15:52:53	U	SMPL	
I3575-07 - 1	HR-06-060817-C	3124	0.6541	-7471A	PPB	12 Jun 2017 15:55:11	U	SMPL	
I3576-03 - 1	TP-OGR1-OGR2	1153	0.2431	-7471A	PPB	12 Jun 2017 15:57:27	U	SMPL	
CCV25 - 1	CCV25	23541	4.9108	-7471A	PPB	12 Jun 2017 15:59:43	U	SMPL	
CCB25 - 1	CCB25	-58	-0.0093	-7471A	PPB	12 Jun 2017 16:02:00	U	SMPL	
PB99649BL - 1	PBS	-63	-0.0104	-7471A	PPB	12 Jun 2017 16:04:16	U	SMPL	
PB99649BS - 1	LCSS	17322	3.6142	-7471A	PPB	12 Jun 2017 16:06:32	U	SMPL	
I3583-01 - 1	RO-2669	771	0.1635	-7471A	PPB	12 Jun 2017 16:08:48	U	SMPL	
I3583-04 - 1	RO-2342	2056	0.4314	-7471A	PPB	12 Jun 2017 16:11:05	U	SMPL	
I3584-01 - 1	DUCT-BANK	1056	0.2229	-7471A	PPB	12 Jun 2017 16:13:22	U	SMPL	
I3591-01 - 1	E2-Q11-ESW	2567	0.5380	-7471A	PPB	12 Jun 2017 16:15:39	U	SMPL	
I3591-02 - 1	E2-R10-BASE	11202	2.3382	-7471A	PPB	12 Jun 2017 16:17:56	U	SMPL	
I3591-03 - 1	E2-R8-BASE	61393	12.8024	-7471A	PPB	12 Jun 2017 16:20:14	U	SMPL	
I3591-04 - 1	E1-R9-BASE	440938	HIGH	-7471A	PPB	12 Jun 2017 16:22:31	X		
I3592-01 - 1	SU-03-060917	2669	0.5592	-7471A	PPB	12 Jun 2017 16:24:21	U	SMPL	
CCV26 - 1	CCV26	23415	4.8845	-7471A	PPB	12 Jun 2017 16:26:40	U	SMPL	
CCB26 - 1	CCB26	-85	-0.0150	-7471A	PPB	12 Jun 2017 16:28:56	U	SMPL	
I3594-01 - 1	EO-01-060917-A	6164	1.2879	-7471A	PPB	12 Jun 2017 16:31:12	U	SMPL	
I3594-01DUP - 1	EO-01-060917-ADUP	6159	1.2868	-7471A	PPB	12 Jun 2017 16:33:29	U	SMPL	
I3594-01MS - 1	EO-01-060917-AMS	22032	4.5962	-7471A	PPB	12 Jun 2017 16:43:46	U	SMPL	
I3594-01MSD - 1	EO-01-060917-AMSD	23019	4.8019	-7471A	PPB	12 Jun 2017 16:46:02	U	SMPL	
I3594-04 - 1	EO-01-060917-B	1360	0.2863	-7471A	PPB	12 Jun 2017 16:48:19	U	SMPL	
I3594-07 - 1	EO-01-060917-C	2082	0.4368	-7471A	PPB	12 Jun 2017 16:50:36	U	SMPL	
I3597-01 - 1	341-COMP	8717	1.8201	-7471A	PPB	12 Jun 2017 16:52:53	U	SMPL	
I3598-01 - 1	VNJ-204-COMP	11115	2.3201	-7471A	PPB	12 Jun 2017 16:55:11	U	SMPL	
I3598-05 - 1	VNJ-226-COMP	1552	0.3263	-7471A	PPB	12 Jun 2017 16:57:29	U	SMPL	
I3591-04DLX10 - 1		69777	14.5504	-7471A	PPB	12 Jun 2017 16:59:45	U	SMPL	
CCV27 - 1	CCV27	23807	4.9662	-7471A	PPB	12 Jun 2017 17:02:02	U	SMPL	
CCB27 - 1	CCB27	-64	-0.0106	-7471A	PPB	12 Jun 2017 17:04:18	U	SMPL	
I3575-01LX5 - 1		129	0.0297	-7471A	PPB	12 Jun 2017 17:06:35	U	SMPL	
I3575-01A - 1		18478	3.8552	-7471A	PPB	12 Jun 2017 17:08:51	U	SMPL	
I3594-01LX5 - 1		1147	0.2419	-7471A	PPB	12 Jun 2017 17:11:08	U	SMPL	
I3594-01A - 1		23772	4.9589	-7471A	PPB	12 Jun 2017 17:13:24	U	SMPL	
I3591-03DLX5 - 1		12432	2.5947	-7471A	PPB	12 Jun 2017 17:15:41	U	SMPL	
I3591-04DL2X50 - 1		13939	2.9089	-7471A	PPB	12 Jun 2017 17:17:57	U	SMPL	
CCV28 - 1	CCV28	23913	4.9883	-7471A	PPB	12 Jun 2017 17:20:14	U	SMPL	
CCB28 - 1	CCB28	-59	-0.0095	-7471A	PPB	12 Jun 2017 17:22:31	U	SMPL	