



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB88023

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	06/12/2017	10:10
2	I3561-01		1	>100.0	>212.0	06/12/2017	10:45
3	I3561-01DUP		1	>100.0	>212.0	06/12/2017	11:10
4	I3561-03		1	>100.0	>212.0	06/12/2017	11:45

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 886023

WorkList Name : FP61217

WorkList ID : 99459

Date : 6/12/2017 9:27:32 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3561-01	Flash Point	Cool 4 deg C	SEAC01	L22	PURGE-WATER-NRL-65	06/07/2017	1010A
	Water	I3561-03	Flash Point	Cool 4 deg C	SEAC01	L22	PURGE-WATER-NRL-74	06/07/2017	1010A

Date/Time 06/12/17 9:30 AM
 Received by: KS
 Relinquished by: KS

Date/Time 06/12/17 11:50 AM
 Received by: KS
 Relinquished by: KS