

CB88024

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

6/12/2017 16:02

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.964	0.0	0.105	
ICB1	-0.001	0.0	0.002	
CCV1	0.966	0.0	0.105	
CCB1	-0.001	0.0	0.002	
PB99690BL	0.001	0.0	0.002	
PB99690BS	0.967	0.0	0.105	
I3615-01	-0.001	0.0	0.002	
I3615-01DUP	-0.001	0.0	0.002	
I3615-01MS	0.968	0.0	0.105	
I3615-01MSD	0.966	0.0	0.105	
I3615-05	-0.004	0.0	0.002	
CCV2	0.967	0.0	0.105	
CCB2	0.001	0.0	0.002	

N	13
Mean	0.445
SD	0.5019
CV%	112.67

A12

CB 88024

Aquakem v. 7.2AQ1

Results from time period:

Mon Jun 12 15:28:40 2017

Mon Jun 12 16:02:11 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0512	mg/l	6/12/2017 15:28:40	
0.1PPM	A	Phenolics-	P	0.1035	mg/l	6/12/2017 15:28:41	
0.25PPM	A	Phenolics-	P	0.2521	mg/l	6/12/2017 15:28:42	
0.50PPM	A	Phenolics-	P	0.4994	mg/l	6/12/2017 15:28:43	
1.0PPM	A	Phenolics-	P	0.9882	mg/l	6/12/2017 15:28:44	
2.0PPM	A	Phenolics-	P	2.0056	mg/l	6/12/2017 15:28:45	
ICV1	S	Phenolics-	P	0.9638	mg/l	6/12/2017 15:58:12	
ICB1	S	Phenolics-	P	-0.0012	mg/l	6/12/2017 15:58:13	
CCV1	S	Phenolics-	P	0.9657	mg/l	6/12/2017 15:58:14	
CCB1	S	Phenolics-	P	-0.0013	mg/l	6/12/2017 15:58:15	
PB99690BL	S	Phenolics-	P	0.0006	mg/l	6/12/2017 15:58:16	
PB99690BS	S	Phenolics-	P	0.9666	mg/l	6/12/2017 15:58:17	
I3615-01	S	Phenolics-	P	-0.0012	mg/l	6/12/2017 15:58:18	
I3615-01DUP	S	Phenolics-	P	-0.0006	mg/l	6/12/2017 15:58:19	
I3615-01MS	S	Phenolics-	P	0.9683	mg/l	6/12/2017 15:58:20	
I3615-01MSD	S	Phenolics-	P	0.9656	mg/l	6/12/2017 15:58:21	
I3615-05	S	Phenolics-	P	-0.0039	mg/l	6/12/2017 15:58:22	
CCV2	S	Phenolics-	P	0.9673	mg/l	6/12/2017 15:58:23	
CCB2	S	Phenolics-	P	0.0005	mg/l	6/12/2017 16:02:10	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

6/12/2017 15:29


Test Phenolics-

Accepted 6/12/2017 15:29

Factor 9.389
Bias 0.002

Coeff. of det. 0.999932

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0512	0.0500	
2	Phenol-2	0.013	0.1035	0.1000	
3	Phenol-2	0.029	0.2521	0.2500	
4	Phenol-2	0.055	0.4994	0.5000	
5	Phenol-2	0.107	0.9882	1.0000	
6	Phenol-2	0.216	2.0056	2.0000	