


7471A

LB 88094
INSTRUMENT ID: CV2

Linear

 μ Abs.:

44764



A= 0.0000e+000

B= 2.2459e-004

C= -5.4701e-002

Rho= 0.9999929

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	202	0.000	202				
0.2	0.200	0.189	-0.011	1084	0.0 %	1084				
2.5	2.500	2.524	0.024	11482	0.0 %	11482				
5.0	5.000	5.012	0.012	22560	0.0 %	22560				
7.5	7.500	7.485	-0.015	33572	0.0 %	33572				
10.0	10.000	9.999	-0.001	44764	0.0 %	44764				

LB88034 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 13 Jun 2017 09:01:34

Sample ID	Extended ID	μ Abs	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	202	-	0.0000	7471A	PPB	13 Jun 2017 09:14:37	S	Std
0.2 - 1	50.2	1084	-	0.2000	7471A	PPB	13 Jun 2017 09:16:55	S	Std
2.5 - 1	52.5	11482	-	2.5000	7471A	PPB	13 Jun 2017 09:19:12	S	Std
5.0 - 1	55	22560	-	5.0000	7471A	PPB	13 Jun 2017 09:21:30	S	Std
7.5 - 1	57.5	33572	-	7.5000	7471A	PPB	13 Jun 2017 09:23:47	S	Std
10.0 - 1	510	44764	-	10.0000	7471A	PPB	13 Jun 2017 09:26:03	S	Std
ICV95 - 1	ICV95	16964	3.7553	-	7471A	PPB	13 Jun 2017 09:28:49	U	SMPL
ICB95 - 1	ICB95	200	-0.0098	-	7471A	PPB	13 Jun 2017 09:31:05	U	SMPL
CCV31 - 1	CCV31	22174	4.9254	-	7471A	PPB	13 Jun 2017 09:33:20	U	SMPL
CCB31 - 1	CCB31	6	-0.0534	-	7471A	PPB	13 Jun 2017 09:35:36	U	SMPL
CRA - 1	CRA	876	0.1420	-	7471A	PPB	13 Jun 2017 09:40:30	U	SMPL
HighStd - 1	HighStd	43607	9.7392	-	7471A	PPB	13 Jun 2017 09:42:46	U	SMPL
ChkStd - 1	ChkStd	30825	6.8684	-	7471A	PPB	13 Jun 2017 09:45:02	U	SMPL
PB99708BL - 1	PBS	170	-0.0165	-	7471A	PPB	13 Jun 2017 09:47:18	U	SMPL
PB99708BS - 1	LCSS	15242	3.3686	-	7471A	PPB	13 Jun 2017 09:49:35	U	SMPL
I3615-01 - 1	HA4-WCA	7006	1.5188	-	7471A	PPB	13 Jun 2017 09:51:52	U	SMPL
I3615-05 - 1	HA4-WCB	807	0.1265	-	7471A	PPB	13 Jun 2017 09:54:09	U	SMPL
I3624-01 - 1	NB-RO6151-060917	4273	0.9050	-	7471A	PPB	13 Jun 2017 09:56:27	U	SMPL
I3624-04 - 1	NB-RO6822-060917	5628	1.2093	-	7471A	PPB	13 Jun 2017 09:58:45	U	SMPL
I3628-02 - 1	COMPOSITE-1	22153	4.9207	-	7471A	PPB	13 Jun 2017 10:01:01	U	SMPL
CCV32 - 1	CCV32	22116	4.9124	-	7471A	PPB	13 Jun 2017 10:03:17	U	SMPL
CCB32 - 1	CCB32	-17	-0.0585	-	7471A	PPB	13 Jun 2017 10:05:33	U	SMPL
I3629-01 - 1	CLINTON-AVE-SP-COMP-1	2258	0.4524	-	7471A	PPB	13 Jun 2017 10:07:49	U	SMPL
I3629-05 - 1	CLINTON-AVE-SP-COMP-2	1974	0.3886	-	7471A	PPB	13 Jun 2017 10:10:05	U	SMPL
I3629-05DUP - 1	CLINTON-AVE-SP-COMP-2DUP	2083	0.4131	-	7471A	PPB	13 Jun 2017 10:12:21	U	SMPL
I3629-05MS - 1	CLINTON-AVE-SP-COMP-2MS	19219	4.2618	-	7471A	PPB	13 Jun 2017 10:14:37	U	SMPL
I3629-05MSD - 1	CLINTON-AVE-SP-COMP-2MSD	17628	3.9044	-	7471A	PPB	13 Jun 2017 10:16:54	U	SMPL
I3637-03DLX10 - 1	5434-5397-5398-COMP	4	-0.0538	-	7471A	PPB	13 Jun 2017 10:19:11	U	SMPL
I3637-04DLX10 - 1	5394-5395-5396-COMP	-120	-0.0817	-	7471A	PPB	13 Jun 2017 10:21:29	U	SMPL
I3638-01DLX10 - 1	17K2719	-136	-0.0852	-	7471A	PPB	13 Jun 2017 10:23:46	U	SMPL
I3638-02DLX10 - 1	1705305428	-88	-0.0745	-	7471A	PPB	13 Jun 2017 10:26:04	U	SMPL
I3639-03DLX10 - 1	1706015445	-78	-0.0722	-	7471A	PPB	13 Jun 2017 10:28:20	U	SMPL
CCV33 - 1	CCV33	21932	4.8711	-	7471A	PPB	13 Jun 2017 10:30:36	U	SMPL
CCB33 - 1	CCB33	-15	-0.0581	-	7471A	PPB	13 Jun 2017 10:32:52	U	SMPL
I3639-04DLX10 - 1	1706015446	-136	-0.0852	-	7471A	PPB	13 Jun 2017 10:35:09	U	SMPL
I3639-05DLX10 - 1	1705265422	-98	-0.0767	-	7471A	PPB	13 Jun 2017 10:37:25	U	SMPL
I3629-05LX5 - 1		364	0.0271	-	7471A	PPB	13 Jun 2017 10:39:41	U	SMPL
I3629-05A - 1		19135	4.2429	-	7471A	PPB	13 Jun 2017 10:41:58	U	SMPL
CCV34 - 1	CCV34	22101	4.9090	-	7471A	PPB	13 Jun 2017 10:44:14	U	SMPL
CCB34 - 1	CCB34	46	-0.0444	-	7471A	PPB	13 Jun 2017 10:46:31	U	SMPL