



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88036

Temp.: 19.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/09/2017	14:00
2	CAL2	1	Water			7.01	06/09/2017	14:01
3	CAL3	1	Water			10.02	06/09/2017	14:03
4	ICV	1	Water			6.99	06/09/2017	14:06
5	CCV1	1	Water			2.01	06/09/2017	14:09
6	I3514-01	1	Water			7.47	06/09/2017	14:10
7	I3514-01DUP	1	Water			7.48	06/09/2017	14:11
8	CCV2	1	Water			12.02	06/09/2017	14:12

LB88036

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3514

WorkList ID : 99514

Date : 6/9/2017 13:30:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3514-01-17	pH	Cool 4 deg C	WAYN02	K11	WASTEWATER	06/09/2017	SM 4500-PH B

Date/Time 06/09/17 13:30 PM
Received by: JP
Relinquished by: JP

Date/Time 06/09/17 3:50 PM
Received by: JP
Relinquished by: JP