


7470A

LB88043
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36177



A= 0.0000e+000

B= 2.7473e-004

C= -3.2514e-002

Rho= 0.9998745

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.039	-0.039	-25	0.000	-25				
0.2	0.200	0.184	-0.016	789	0.0 %	789				
2.5	2.500	2.519	0.019	9289	0.0 %	9289				
5.0	5.000	5.040	0.040	18463	0.0 %	18463				
7.5	7.500	7.589	0.089	27743	0.0 %	27743				
10.0	10.000	9.906	-0.094	36177	0.0 %	36177				

LB88043 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Jun 2017 12:22:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-25	-	0.00007470A	PPB		13 Jun 2017 12:26:20	S	Std
0.2 - 1	50.2	789	-	0.20007470A	PPB		13 Jun 2017 12:28:29	S	Std
2.5 - 1	52.5	9289	-	2.50007470A	PPB		13 Jun 2017 12:30:37	S	Std
5.0 - 1	55	18463	-	5.00007470A	PPB		13 Jun 2017 12:32:46	S	Std
7.5 - 1	57.5	27743	-	7.50007470A	PPB		13 Jun 2017 12:34:54	S	Std
10.0 - 1	510	36177	-	10.00007470A	PPB		13 Jun 2017 12:39:23	S	Std
ICV96 - 1	ICV96	14580	3.9731	-7470A	PPB		13 Jun 2017 12:41:54	U	SMPL
ICB96 - 1	ICB96	-73	-0.0526	-7470A	PPB		13 Jun 2017 12:44:01	U	SMPL
CCV35 - 1	CCV35	18634	5.0868	-7470A	PPB		13 Jun 2017 12:46:07	U	SMPL
CCB35 - 1	CCB35	-96	-0.0589	-7470A	PPB		13 Jun 2017 12:48:14	U	SMPL
CRA - 1	CRA	749	0.1733	-7470A	PPB		13 Jun 2017 12:50:20	U	SMPL
HighStd - 1	HighStd	36354	9.9551	-7470A	PPB		13 Jun 2017 12:52:27	U	SMPL
ChkStd - 1	ChkStd	25854	7.0704	-7470A	PPB		13 Jun 2017 12:54:34	U	SMPL
PB99691BL - 1	PBW	-83	-0.0553	-7470A	PPB		13 Jun 2017 12:56:41	U	SMPL
PB99691BS - 1	LCSW	14000	3.8137	-7470A	PPB		13 Jun 2017 12:58:48	U	SMPL
PB99691TB - 1	PB99634TB	-88	-0.0567	-7470A	PPB		13 Jun 2017 13:00:56	U	SMPL
I3576-06 - 1	TP-OGR1-OGR2	-40	-0.0435	-7470A	PPB		13 Jun 2017 13:03:04	U	SMPL
I3583-03 - 1	RO-2669	-34	-0.0419	-7470A	PPB		13 Jun 2017 13:05:12	U	SMPL
I3583-06 - 1	RO-2342	-32	-0.0413	-7470A	PPB		13 Jun 2017 13:07:21	U	SMPL
I3584-03 - 1	DUCT-BANK	-30	-0.0408	-7470A	PPB		13 Jun 2017 13:09:28	U	SMPL
CCV36 - 1	CCV36	18812	5.1357	-7470A	PPB		13 Jun 2017 13:11:35	U	SMPL
CCB36 - 1	CCB36	-73	-0.0526	-7470A	PPB		13 Jun 2017 13:13:42	U	SMPL
I3586-02 - 1	CORE-SAMPLE-NRL-65	-50	-0.0463	-7470A	PPB		13 Jun 2017 13:15:48	U	SMPL
I3587-02 - 1	CORE-SAMPLE-NRL-74	-44	-0.0446	-7470A	PPB		13 Jun 2017 13:17:55	U	SMPL
I3592-03 - 1	SU-03-060917	-45	-0.0449	-7470A	PPB		13 Jun 2017 13:20:02	U	SMPL
I3592-03DUP - 1	SU-03-060917DUP	-35	-0.0421	-7470A	PPB		13 Jun 2017 13:22:09	U	SMPL
I3592-03MS - 1	SU-03-060917MS	13828	3.7665	-7470A	PPB		13 Jun 2017 13:24:17	U	SMPL
I3592-03MSD - 1	SU-03-060917MSD	14010	3.8165	-7470A	PPB		13 Jun 2017 13:26:24	U	SMPL
I3594-03 - 1	EO-01-060917-A	-103	-0.0608	-7470A	PPB		13 Jun 2017 13:28:33	U	SMPL
I3594-06 - 1	EO-01-060917-B	-58	-0.0484	-7470A	PPB		13 Jun 2017 13:30:41	U	SMPL
I3594-09 - 1	EO-01-060917-C	-46	-0.0452	-7470A	PPB		13 Jun 2017 13:32:50	U	SMPL
I3597-02 - 1	341-COMP	-33	-0.0416	-7470A	PPB		13 Jun 2017 13:34:57	U	SMPL
CCV37 - 1	CCV37	18701	5.1052	-7470A	PPB		13 Jun 2017 13:37:03	U	SMPL
CCB37 - 1	CCB37	-64	-0.0501	-7470A	PPB		13 Jun 2017 13:39:10	U	SMPL
I3598-03 - 1	VNJ-204-COMP	-20	-0.0380	-7470A	PPB		13 Jun 2017 13:41:17	U	SMPL
I3598-07 - 1	VNJ-226-COMP	25	-0.0256	-7470A	PPB		13 Jun 2017 13:43:24	U	SMPL
PB99707BL - 1	PBW	30	-0.0243	-7470A	PPB		13 Jun 2017 13:45:31	U	SMPL
PB99707BS - 1	LCSW	14162	3.8582	-7470A	PPB		13 Jun 2017 13:47:38	U	SMPL
PB99707TB - 1	PB99652TB	-38	-0.0430	-7470A	PPB		13 Jun 2017 13:49:45	U	SMPL
I3615-02 - 1	HA4-WCA	0	-0.0325	-7470A	PPB		13 Jun 2017 13:51:53	U	SMPL
I3615-06 - 1	HA4-WCB	-12	-0.0358	-7470A	PPB		13 Jun 2017 13:54:00	U	SMPL
I3615-06DUP - 1	HA4-WCBDUP	-5	-0.0339	-7470A	PPB		13 Jun 2017 13:56:08	U	SMPL
I3615-06MS - 1	HA4-WCBMS	13463	3.6662	-7470A	PPB		13 Jun 2017 13:58:17	U	SMPL
I3615-06MSD - 1	HA4-WCBMSD	13723	3.7376	-7470A	PPB		13 Jun 2017 14:00:24	U	SMPL
CCV38 - 1	CCV38	18682	5.1000	-7470A	PPB		13 Jun 2017 14:02:31	U	SMPL
CCB38 - 1	CCB38	-59	-0.0487	-7470A	PPB		13 Jun 2017 14:04:38	U	SMPL
I3624-03 - 1	NB-RO6151-060917	19	-0.0273	-7470A	PPB		13 Jun 2017 14:06:45	U	SMPL
I3624-06 - 1	NB-RO6822-060917	34	-0.0232	-7470A	PPB		13 Jun 2017 14:08:53	U	SMPL
I3592-03LX5 - 1		-13	-0.0361	-7470A	PPB		13 Jun 2017 14:11:00	U	SMPL
I3592-03A - 1		14215	3.8728	-7470A	PPB		13 Jun 2017 14:13:07	U	SMPL
I3615-06LX5 - 1		-83	-0.0553	-7470A	PPB		13 Jun 2017 14:15:15	U	SMPL
I3615-06A - 1		14314	3.9000	-7470A	PPB		13 Jun 2017 14:17:22	U	SMPL
CCV39 - 1	CCV39	18181	4.9624	-7470A	PPB		13 Jun 2017 14:19:31	U	SMPL
CCB39 - 1	CCB39	-72	-0.0523	-7470A	PPB		13 Jun 2017 14:21:39	U	SMPL